



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042923\
Data File : VU053964.D
Acq On : 28 Apr 2023 11:03
Operator : JC/MD
Sample : BFB181
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB181

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Title : VOC Analysis
Last Update : Fri Apr 14 01:54:10 2023



AutoFind: Scans 2888, 2889, 2890; Background Corrected with Scan 2877

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.0	5785	PASS
75	95	30	80	49.0	17739	PASS
95	95	100	100	100.0	36205	PASS
96	95	5	9	6.0	2157	PASS
173	174	0.00	2	0.9	265	PASS
174	95	50	120	85.0	30784	PASS
175	174	5	9	7.3	2234	PASS
176	174	95	101	100.1	30813	PASS
177	176	5	9	5.9	1810	PASS

m/z	Abundance
44.10	196.0

Instrument :
MSVOA_U
ClientSampleId :
BFB181

m/z	Abundance
36.00	307.0
37.10	1617.0
38.10	1474.0
39.00	538.0
44.00	464.0
45.10	302.0
47.10	292.0
47.90	245.0
49.00	1282.0
50.10	5755.0
51.10	1813.0
55.20	116.0
56.10	589.0
57.10	904.0
60.00	408.0
61.10	1747.0
62.10	1736.0
63.00	1287.0
64.10	134.0
65.10	279.0
68.10	3505.0
69.00	3314.0
70.00	262.0
72.30	182.0
73.10	1649.0
74.10	6083.0
75.10	17936.0
76.10	1597.0
77.10	221.0
77.90	254.0
79.00	1486.0
80.00	290.0
81.00	1691.0
82.00	278.0
87.00	1425.0
88.00	1662.0
91.10	156.0
92.00	1202.0
93.10	1589.0
94.10	4744.0
95.10	36224.0
96.10	2065.0
103.80	201.0
106.10	202.0
116.00	180.0
117.00	363.0
117.90	196.0
118.90	315.0
128.00	110.0
130.00	163.0
137.00	108.0
141.10	370.0
143.00	536.0
155.00	122.0
172.00	379.0
173.00	251.0
174.00	29568.0
175.00	2356.0
176.00	29456.0
177.00	1765.0

Instrument :
MSVOA_U
ClientSampleId :
BFB181

m/z	Abundance
36.00	302.0
37.10	1406.0
38.10	1559.0
39.10	546.0
44.00	512.0
45.10	321.0
47.00	312.0
47.90	163.0
49.00	1160.0
50.10	5891.0
51.10	1840.0
55.20	102.0
56.10	546.0
57.10	799.0
60.00	464.0
61.10	1808.0
62.10	1742.0
63.00	1221.0
64.10	119.0
65.00	342.0
67.10	105.0
68.10	3749.0
69.10	3293.0
70.00	291.0
72.20	206.0
73.10	1671.0
74.10	6321.0
75.10	18272.0
76.10	1748.0
77.10	221.0
77.90	281.0
79.00	1529.0
79.90	309.0
81.00	1742.0
82.00	311.0
87.00	1633.0
88.10	1538.0
91.00	172.0
92.00	1215.0
93.10	1695.0
94.10	4735.0
95.10	37200.0
96.10	2257.0
103.90	233.0
106.10	221.0
116.00	161.0
117.00	430.0
117.90	165.0
118.80	279.0
128.00	104.0
130.00	153.0
137.00	126.0
141.00	377.0
142.90	555.0
144.80	112.0
155.00	153.0
172.00	413.0
173.00	271.0
174.00	31792.0
175.00	2321.0
176.00	31864.0
177.00	1828.0

Instrument :
MSVOA_U
ClientSampleId :
BFB181

m/z	Abundance
36.00	262.0
37.10	1183.0
38.10	1480.0
39.10	550.0
44.00	544.0
45.00	301.0
47.00	323.0
48.00	105.0
49.00	1029.0
50.10	5709.0
51.10	1718.0
56.00	490.0
57.10	700.0
60.00	393.0
61.10	1683.0
62.00	1555.0
63.00	1185.0
65.00	318.0
67.10	112.0
68.10	3477.0
69.10	3050.0
70.00	268.0
72.20	162.0
73.00	1522.0
74.10	6146.0
75.10	17008.0
76.10	1659.0
77.10	224.0
77.90	274.0
79.00	1459.0
79.90	367.0
81.00	1749.0
82.00	288.0
87.00	1677.0
88.10	1319.0
91.00	184.0
92.00	1144.0
93.10	1564.0
94.10	4301.0
95.10	35192.0
96.10	2150.0
103.90	204.0
106.10	186.0
115.90	101.0
117.00	391.0
118.00	112.0
118.80	228.0
130.00	152.0
137.00	162.0
140.90	400.0
142.90	522.0
155.00	148.0
171.90	410.0
173.00	274.0
174.00	30992.0
175.00	2026.0
176.00	31120.0
177.00	1837.0

Instrument :
MSVOA_U
ClientSampleId :
BFB181