

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU043020\
 Data File : VU037977.D
 Acq On : 01 May 2020 02:21
 Operator : JC/MD
 Sample : L2395-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: May 01 06:16:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM043020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 01 06:13:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	707990	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	691916	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	314319	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	234167	37.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.20%
7) Chloroethane-d5	1.93	69	210487	41.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.10%
11) 1,1-Dichloroethene-d2	2.59	63	302291	29.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.34%#
21) 2-Butanone-d5	4.66	46	422879	86.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.88%
24) Chloroform-d	5.10	84	391862	40.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.28%
26) 1,2-Dichloroethane-d4	5.74	65	297180	44.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.74%
32) Benzene-d6	5.76	84	892865	43.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.56%
36) 1,2-Dichloropropane-d6	6.72	67	301715	44.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.24%
41) Toluene-d8	7.92	98	796825	42.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.18%
43) trans-1,3-Dichloropropene-	8.20	79	131225	41.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.76%
47) 2-Hexanone-d5	8.65	63	271772	92.41	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.41%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	417983	42.90	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	291944	44.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.84%

Target Compounds


					Ovalue
25) Chloroform	5.13	83	28033	2.928 ug/L	92
48) 2-Hexanone	8.71	43	37204	4.818 ug/L #	83
59) 1,2,3-Trichloropropane	11.82	75	44249	5.518 ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU043020\
Data File : VU037977.D
Acq On : 01 May 2020 02:21
Operator : JC/MD
Sample : L2395-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: May 01 06:16:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM043020WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 01 06:13:06 2020
Response via : Initial Calibration









#25
Chloroform
Concen: 2.928 ug/L
RT: 5.13 min Scan# 1178
Delta R.T. 0.00 min
Lab File: VU037977.D
Acq: 01 May 2020 02:21




Instrument :
MSVOA_U
ClientSampled :


Tot Ion: 83 Resp: 28033
Ion Ratio Lower Upper
83 100
85 70.5 45.0 83.6



#48
2-Hexanone
Concen: 4.818 ug/L
RT: 8.71 min Scan# 2292
Delta R.T. 0.00 min
Lab File: VU037977.D
Acq: 01 May 2020 02:21

Tgt Ion: 43 Resp: 37204
Ion Ratio Lower Upper
43 100
58 48.9 46.4 69.6
57 3.9 16.4 24.6#
100 10.1 9.5 14.3





#59

1,2,3-Trichloropropane

Concen: 5.518 ug/L

RT: 11.82 min Scan# 3260

Delta R.T. 0.98 min

Lab File: VU037977.D

Acq: 01 May 2020 02:21

Instrument :

MSVOA_U

ClientSampleId :

Tot Ion: 75 Resp: 44249

Ion Ratio Lower Upper

75 100

77 34.8 25.4 38.2

