

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU050620\
 Data File : VU038025.D
 Acq On : 06 May 2020 16:58
 Operator : JC/MD
 Sample : L2488-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: May 07 10:35:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM043020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 07 01:04:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	715823	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	705372	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	326269	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	231449	36.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.52%
7) Chloroethane-d5	1.93	69	213769	41.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.48%
11) 1,1-Dichloroethene-d2	2.59	63	304772	29.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.18%#
21) 2-Butanone-d5	4.66	46	481793	97.90	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.90%
24) Chloroform-d	5.10	84	408892	41.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.88%
26) 1,2-Dichloroethane-d4	5.74	65	300907	44.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.86%
32) Benzene-d6	5.76	84	895850	43.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.18%
36) 1,2-Dichloropropane-d6	6.72	67	301211	43.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.38%
41) Toluene-d8	7.92	98	806527	42.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.58%
43) trans-1,3-Dichloropropene-	8.20	79	133038	41.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.32%
47) 2-Hexanone-d5	8.65	63	277482	92.55	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.55%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	437917	44.09	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.18%
64) 1,2-Dichlorobenzene-d4	12.20	152	294994	43.24	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU050620\
Data File : VU038025.D
Acq On : 06 May 2020 16:58
Operator : JC/MD
Sample : L2488-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: May 07 10:35:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM043020WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 07 01:04:11 2020
Response via : Initial Calibration

