

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
 Data File : VU043576.D
 Acq On : 06 May 2021 12:31
 Operator : SY/MD
 Sample : M2251-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW8-EB-20210429

Quant Time: May 07 02:03:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	117904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	210111	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	203565	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	93353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	122166	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	
35) Dibromofluoromethane	5.295	113	82617	49.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.60%	
50) Toluene-d8	7.906	98	289928	48.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.60%	
62) 4-Bromofluorobenzene	10.639	95	99902	44.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.40%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
Data File : VU043576.D
Acq On : 06 May 2021 12:31
Operator : SY/MD
Sample : M2251-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW8-EB-20210429

Quant Time: May 07 02:03:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

