

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
 Data File : VU043584.D
 Acq On : 06 May 2021 15:36
 Operator : SY/MD
 Sample : M2243-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 07 02:05:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	108393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.257	114	196368	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.424	117	191575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	83968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	116502	53.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.18%	
35) Dibromofluoromethane	5.295	113	79058	50.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.96%	
50) Toluene-d8	7.906	98	273634	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
62) 4-Bromofluorobenzene	10.639	95	94756	44.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
Data File : VU043584.D
Acq On : 06 May 2021 15:36
Operator : SY/MD
Sample : M2243-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 07 02:05:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

