

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU050621\  
 Data File : VU043586.D  
 Acq On : 06 May 2021 16:23  
 Operator : SY/MD  
 Sample : M2313-02  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 210505001-01-TRIP-BLANK

Quant Time: May 07 02:05:37 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U050621W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 06 07:26:23 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| -----                       |        |      |          |       |         |          |
| Internal Standards          |        |      |          |       |         |          |
| 1) Pentafluorobenzene       | 5.382  | 168  | 110618   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.256  | 114  | 197660   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 9.423  | 117  | 196803   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.819 | 152  | 87431    | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |        |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 5.710  | 65   | 117409   | 52.43 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 104.86% |          |
| 35) Dibromofluoromethane    | 5.298  | 113  | 80340    | 50.97 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 101.94% |          |
| 50) Toluene-d8              | 7.906  | 98   | 279858   | 50.07 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 100.14% |          |
| 62) 4-Bromofluorobenzene    | 10.639 | 95   | 95767    | 45.04 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 90.08%  |          |
| Target Compounds            |        |      |          |       |         |          |
| 16) Acetone                 | 2.633  | 43   | 3844     | 3.09  | ug/l    | 91       |
| 30) Chloroform              | 5.102  | 83   | 1075     | 0.35  | ug/l #  | 65       |
| -----                       |        |      |          |       |         |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU050621\  
Data File : VU043586.D  
Acq On : 06 May 2021 16:23  
Operator : SY/MD  
Sample : M2313-02  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
210505001-01-TRIP-BLANK

Quant Time: May 07 02:05:37 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U050621W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 06 07:26:23 2021  
Response via : Initial Calibration

