

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
 Data File : VU043588.D
 Acq On : 06 May 2021 17:09
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK

Quant Time: May 07 02:06:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	115787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	203911	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	198596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	89883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	120783	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
35) Dibromofluoromethane	5.298	113	81723	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
50) Toluene-d8	7.906	98	284763	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	10.639	95	99292	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
Target Compounds						
						Qvalue
16) Acetone	2.633	43	1632	1.25	ug/l	90
20) Methylene Chloride	3.057	84	1953	1.07	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
Data File : VU043588.D
Acq On : 06 May 2021 17:09
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK

Quant Time: May 07 02:06:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

