

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
 Data File : VU043591.D
 Acq On : 06 May 2021 18:17
 Operator : SY/MD
 Sample : M2267-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CITY-WATER

Quant Time: May 07 02:06:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	111773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	200509	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	199429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	88950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	120212	53.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.26%	
35) Dibromofluoromethane	5.298	113	82412	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
50) Toluene-d8	7.906	98	280622	49.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.00%	
62) 4-Bromofluorobenzene	10.639	95	99554	46.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.30%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	9450	7.52	ug/l	99
30) Chloroform	5.096	83	33383	10.74	ug/l	99
47) Bromodichloromethane	7.112	83	24176	9.42	ug/l	99
60) Dibromochloromethane	8.816	129	13842	7.57	ug/l	99
71) Bromoform	10.298	173	1743	4.35	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
Data File : VU043591.D
Acq On : 06 May 2021 18:17
Operator : SY/MD
Sample : M2267-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CITY-WATER

Quant Time: May 07 02:06:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

