

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
 Data File : VU043596.D
 Acq On : 06 May 2021 20:13
 Operator : SY/MD
 Sample : M2307-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-18

Quant Time: May 07 02:08:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 06 07:26:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	116707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	210411	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	206592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	91944	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	122133	51.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.40%	
35) Dibromofluoromethane	5.298	113	83539	49.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.56%	
50) Toluene-d8	7.906	98	294469	49.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.98%	
62) 4-Bromofluorobenzene	10.639	95	103586	45.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.52%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	2022	1.54	ug/l	94
95) Naphthalene	14.095	128	13692	3.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050621\
Data File : VU043596.D
Acq On : 06 May 2021 20:13
Operator : SY/MD
Sample : M2307-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-18

Quant Time: May 07 02:08:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U050621W.M
Quant Title : SW846 8260
QLast Update : Thu May 06 07:26:23 2021
Response via : Initial Calibration

