

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050724\
 Data File : VU058845.D
 Acq On : 07 May 2024 16:35
 Operator : MD/SY
 Sample : P2392-03DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAQ2DL

Quant Time: May 08 05:11:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 08 05:10:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	58284	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	55911	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	25110	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	15147	3.304	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	66.000%	
7) Chloroethane-d5	1.904	69	12359	3.578	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	71.600%	
11) 1,1-Dichloroethene-d2	2.557	65	9362	4.056	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	81.200%	
20) 2-Butanone-d5	4.608	46	38502	47.800	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.600%	
24) Chloroform-d	5.052	84	49616	5.294	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.800%	
26) 1,2-Dichloroethane-d4	5.692	65	26575	5.850	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	117.000%	
32) Benzene-d6	5.718	84	83579	4.696	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	6.682	67	23610	4.448	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.000%	
41) Toluene-d8	7.894	98	71461	4.486	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.800%	
43) trans-1,3-Dichloroprop...	8.177	79	8907	4.609	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.200%	
46) 2-Hexanone-d5	8.624	63	27172	47.448	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.900%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	16708	5.168	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	24400	5.575	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.600%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.573	96	1203	0.283	ug/L	# 1
16) Methylene chloride	3.033	84	3023	0.662	ug/L	90
22) cis-1,2-Dichloroethene	4.656	96	6238	1.336	ug/L	94
25) Chloroform	5.078	83	7429	0.794	ug/L	96
34) Trichloroethene	6.537	95	2232	0.424	ug/L	97
47) Tetrachloroethene	8.547	164	28152	7.621	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050724\
Data File : VU058845.D
Acq On : 07 May 2024 16:35
Operator : MD/SY
Sample : P2392-03DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAQ2DL

Quant Time: May 08 05:11:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 08 05:10:37 2024
Response via : Initial Calibration

