

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050724\
 Data File : VU058850.D
 Acq On : 07 May 2024 19:10
 Operator : MD/SY
 Sample : P2392-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAQ8

Quant Time: May 08 05:12:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 08 05:10:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	54848	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	54063	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	24738	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.589	65	12793	2.965	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.400%
7) Chloroethane-d5	1.904	69	11240	3.457	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.200%
11) 1,1-Dichloroethene-d2	2.557	65	8275	3.810	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.200%
20) 2-Butanone-d5	4.612	46	35928	47.399	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.800%
24) Chloroform-d	5.049	84	44835	5.083	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
26) 1,2-Dichloroethane-d4	5.692	65	24255	5.674	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.400%
32) Benzene-d6	5.718	84	72117	4.191	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
36) 1,2-Dichloropropane-d6	6.682	67	21760	4.240	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.800%
41) Toluene-d8	7.891	98	63944	4.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	8.174	79	8257	4.419	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.400%
46) 2-Hexanone-d5	8.627	63	24630	44.480	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.960%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	14358	4.593	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	23004	5.335	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	4.657	96	782	0.178	ug/L #	85
25) Chloroform	5.075	83	19979	2.269	ug/L	100
34) Trichloroethene	6.541	95	2451	0.482	ug/L	97
38) Bromodichloromethane	7.100	83	911	0.151	ug/L #	97
47) Tetrachloroethene	8.544	164	9867	2.762	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050724\
Data File : VU058850.D
Acq On : 07 May 2024 19:10
Operator : MD/SY
Sample : P2392-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAQ8

Quant Time: May 08 05:12:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 08 05:10:37 2024
Response via : Initial Calibration

