


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050724\
Data File : VU058836.D
Acq On : 07 May 2024 08:43
Operator : MD/SY
Sample : BFB152
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB152

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR042224WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Wed May 08 05:10:37 2024



AutoFind: Scans 2903, 2904, 2905; Background Corrected with Scan 2891

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.1	13753	PASS
75	95	30	80	56.1	38379	PASS
95	95	100	100	100.0	68381	PASS
96	95	5	9	6.3	4300	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	80.5	55053	PASS
175	174	5	9	7.9	4341	PASS
176	174	95	101	97.1	53475	PASS
177	176	5	9	6.4	3449	PASS

m/z	Abundance
44.00	103.0

Instrument :
MSVOA_U
ClientSampleId :
BFB152

m/z	Abundance
35.90	590.0
36.90	3447.0
37.90	2968.0
39.00	1223.0
43.90	616.0
44.90	610.0
45.90	121.0
46.90	685.0
47.90	487.0
48.90	3085.0
50.00	13262.0
51.00	4210.0
51.90	176.0
54.90	249.0
55.90	1273.0
56.90	2146.0
57.90	103.0
59.90	766.0
61.00	3831.0
62.00	3827.0
62.90	3214.0
64.00	489.0
65.00	440.0
66.80	171.0
67.90	7310.0
69.00	7827.0
69.90	504.0
71.90	445.0
73.00	3344.0
74.00	13662.0
75.00	36760.0
76.00	3198.0
76.80	373.0
77.90	198.0
78.80	2881.0
79.90	755.0
80.80	2868.0
81.90	616.0
86.90	2056.0
87.90	1777.0
90.90	288.0
91.90	2125.0
93.00	3270.0
94.00	8277.0
95.00	65648.0
96.00	4082.0
96.90	107.0
103.90	365.0
104.80	194.0
105.80	430.0
110.60	144.0
112.80	115.0
115.80	381.0
116.80	695.0
117.90	393.0
118.90	447.0
127.90	257.0
128.90	191.0
129.80	287.0
134.80	226.0
136.70	178.0
140.90	998.0
142.90	1077.0
144.80	244.0
145.70	173.0
147.70	189.0
154.80	220.0
156.70	142.0
171.00	147.0
172.00	510.0
173.90	51608.0
174.90	4095.0
175.90	49792.0
176.90	3347.0
177.90	124.0

Instrument :
MSVOA_U
ClientSampleId :
BFB152

m/z	Abundance
35.90	677.0
36.90	3559.0
37.90	3333.0
39.00	1231.0
43.90	621.0
44.90	692.0
45.90	100.0
46.90	747.0
47.90	571.0
48.90	3374.0
50.00	14220.0
51.00	4319.0
51.90	197.0
54.90	270.0
55.90	1453.0
56.90	2255.0
59.90	888.0
61.00	4312.0
62.00	4122.0
63.00	3270.0
63.90	485.0
65.00	488.0
66.80	192.0
67.90	7571.0
69.00	8309.0
69.90	591.0
71.90	471.0
73.00	3588.0
74.00	14712.0
75.00	39528.0
76.00	3308.0
76.80	431.0
77.80	212.0
78.80	3080.0
79.90	825.0
80.80	3012.0
81.90	690.0
86.90	2244.0
87.90	2020.0
90.80	336.0
91.90	2259.0
93.00	3540.0
94.00	8883.0
95.00	70616.0
95.90	4408.0
96.90	149.0
103.90	357.0
104.70	185.0
105.80	455.0
110.70	141.0
112.70	136.0
114.90	105.0
115.70	378.0
116.80	683.0
117.90	413.0
118.90	509.0
127.90	313.0
128.90	242.0
129.80	307.0
134.80	273.0
136.70	176.0
140.90	1135.0
142.90	1194.0
144.90	245.0
145.70	171.0
147.70	197.0
154.80	210.0
156.70	201.0
159.00	103.0
171.00	144.0
172.00	594.0
173.90	56872.0
174.90	4507.0
175.90	55184.0
176.90	3564.0
177.90	123.0

Instrument :
MSVOA_U
ClientSampleId :
BFB152

m/z	Abundance
35.90	723.0
37.00	3458.0
38.00	3216.0
39.00	1188.0
43.90	570.0
44.90	683.0
46.90	697.0
47.90	554.0
48.90	3328.0
50.00	13777.0
51.00	4041.0
52.00	184.0
54.90	249.0
55.90	1450.0
56.90	2132.0
60.00	934.0
61.00	4259.0
62.00	4038.0
63.00	3198.0
63.90	429.0
65.00	536.0
66.80	187.0
67.90	7308.0
69.00	8189.0
69.90	640.0
71.90	479.0
73.00	3416.0
74.00	14371.0
75.00	38848.0
76.00	3203.0
76.80	406.0
77.70	220.0
78.90	3055.0
79.90	819.0
80.80	2980.0
81.90	700.0
86.90	2334.0
87.90	2107.0
90.80	333.0
92.00	2142.0
92.90	3412.0
94.00	8942.0
95.00	68880.0
95.90	4410.0
97.00	158.0
103.90	363.0
104.70	133.0
105.90	439.0
112.70	119.0
115.70	343.0
116.80	640.0
117.90	352.0
118.90	506.0
127.90	348.0
128.80	236.0
129.80	322.0
134.80	262.0
136.70	152.0
140.90	1128.0
142.00	116.0
142.90	1245.0
144.90	260.0
145.70	133.0
147.70	196.0
154.80	159.0
156.70	207.0
159.00	129.0
171.00	138.0
172.00	610.0
173.90	56680.0
174.90	4421.0
175.90	55448.0
176.90	3437.0
177.90	100.0

Instrument :
MSVOA_U
ClientSampleId :
BFB152