

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U051920\
 Data File : VU038235.D
 Acq On : 19 May 2020 15:22
 Operator : JC/MD
 Sample : L2657-07DL 10X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSW0DL

Quant Time: May 20 02:57:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 20 01:25:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	356772	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	344808	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	169355	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	121871	3.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.80%
7) Chloroethane-d5	1.93	69	121216	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.80%
11) 1,1-Dichloroethene-d2	2.58	63	156183	2.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.20%#
20) 2-Butanone-d5	4.67	46	495985	61.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.20%
24) Chloroform-d	5.10	84	229289	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	5.74	65	143994	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.76	84	467940	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.72	67	163728	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	7.92	98	395190	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
43) trans-1,3-Dichloropropene-	8.20	79	59801	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.66	63	397385	63.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	126.00%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	145642	5.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.60%
64) 1,2-Dichlorobenzene-d4	12.21	152	161925	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.39	85	3893	0.119 ug/L	92
9) Trichlorofluoromethane	2.15	101	31992	0.812 ug/L	98
12) 1,1-Dichloroethene	2.60	96	10001	0.404 ug/L #	1
22) cis-1,2-Dichloroethene	4.70	96	22097	0.759 ug/L	96
29) 1,1,1-Trichloroethane	5.35	97	2797	0.071 ug/L	98
34) Trichloroethene	6.57	95	227635	8.243 ug/L	97
47) Tetrachloroethene	8.57	164	261152	13.519 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051920\
Data File : VU038235.D
Acq On : 19 May 2020 15:22
Operator : JC/MD
Sample : L2657-07DL 10X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSW0DL

Quant Time: May 20 02:57:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 20 01:25:15 2020
Response via : Initial Calibration

