

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052124\
 Data File : VU058923.D
 Acq On : 21 May 2024 11:36
 Operator : MD/SY
 Sample : VSTD0.0519
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05019

Quant Time: May 22 06:58:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052124.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 22 06:58:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	6813	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	10620	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	3650	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	289	0.051	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	365	0.052	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	374	0.068	ug/L	0.00
8) Toluene-d8	7.903	98	806	0.056	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.745	84	378	0.073	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	462	0.058	ug/L #	60
6) Trichloroethene	6.544	95	342	0.062	ug/L	95
9) 1,2-Dibromoethane	8.924	107	234	0.062	ug/L #	91
12) 1,2,3-Trichloropropane	10.814	75	261	0.075	ug/L	95
13) 1,2-Dibromo-3-chloropr...	13.008	75	57	0.072	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052124\
Data File : VU058923.D
Acq On : 21 May 2024 11:36
Operator : MD/SY
Sample : VSTD0.0519
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05019

Quant Time: May 22 06:58:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052124.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 22 06:58:22 2024
Response via : Initial Calibration

