

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052124\
 Data File : VU058925.D
 Acq On : 21 May 2024 12:30
 Operator : MD/SY
 Sample : VSTD0.521
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5021

Quant Time: May 22 06:59:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052124.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 22 06:58:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	6640	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	10434	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	3566	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	2760	0.503	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	3085	0.448	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	3779	0.699	ug/L	0.00
8) Toluene-d8	7.887	98	7036	0.502	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3462	0.677	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	3882	0.499	ug/L	100
6) Trichloroethene	6.544	95	2945	0.546	ug/L	100
9) 1,2-Dibromoethane	8.924	107	2238	0.600	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	2380	0.698	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	428	0.555	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052124\
Data File : VU058925.D
Acq On : 21 May 2024 12:30
Operator : MD/SY
Sample : VSTD0.521
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5021

Quant Time: May 22 06:59:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052124.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 22 06:58:22 2024
Response via : Initial Calibration

