

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052524\
 Data File : VU059014.D
 Acq On : 25 May 2024 21:29
 Operator : MD/SY
 Sample : P2566-03 1000X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHHS7

Quant Time: May 27 04:17:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 27 04:10:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	64312	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	63465	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	25347	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	62633	92.082	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 184.160%#	
7) Chloroethane-d5	1.888	69	56770	100.426	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 200.860%#	
11) 1,1-Dichloroethene-d2	2.554	63	80095	77.074	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 154.140%#	
21) 2-Butanone-d5	4.608	46	123954	265.299	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 265.300%#	
24) Chloroform-d	5.052	84	124293	106.534	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 213.060%#	
26) 1,2-Dichloroethane-d4	5.692	65	84625	118.601	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 237.200%#	
32) Benzene-d6	5.718	84	245473	113.453	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 226.900%#	
36) 1,2-Dichloropropane-d6	6.682	67	87393	122.289	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 244.580%#	
41) Toluene-d8	7.891	98	196470	101.305	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 202.620%#	
43) trans-1,3-Dichloroprop...	8.174	79	31301	97.543	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 195.080%#	
47) 2-Hexanone-d5	8.624	63	68761	232.641	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 232.640%#	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	122958	114.367	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 228.740%#	
66) 1,2-Dichlorobenzene-d4	12.187	152	62137	118.808	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 237.620%#	
Target Compounds						
					Qvalue	
34) Trichloroethene	6.531	95	327660	647.039	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052524\
Data File : VU059014.D
Acq On : 25 May 2024 21:29
Operator : MD/SY
Sample : P2566-03 1000X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHHS7

Quant Time: May 27 04:17:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 27 04:10:55 2024
Response via : Initial Calibration

