

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU052524\  
 Data File : VU058993.D  
 Acq On : 25 May 2024 12:54  
 Operator : MD/SY  
 Sample : P2409-05  
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA\_U/MEOH  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 MDL-MED-SOIL-QT2-2024-04

Manual Integrations  
 APPROVED

Reviewed By :Mahesh  
 Dadoda

Quant Time: May 27 05:44:27 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM052324WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Mon May 27 04:10:55 2024  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|---------|-------|----------|----------|--------|----------|
| -----                         |         |       |          |          |        |          |
| Internal Standards            |         |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 6.242   | 114   | 124719   | 50.000   | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 9.412   | 117   | 124549   | 50.000   | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.807  | 152   | 54920    | 50.000   | ug/L   | 0.00     |
| System Monitoring Compounds   |         |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.592   | 65    | 60107    | 45.567   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 135 | Recovery | =      | 91.140%  |
| 7) Chloroethane-d5            | 1.875   | 69    | 56026    | 51.106   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 130 | Recovery | =      | 102.220% |
| 11) 1,1-Dichloroethene-d2     | 2.550   | 63    | 79363    | 39.380   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =      | 78.760%  |
| 21) 2-Butanone-d5             | 4.611   | 46    | 113377   | 125.129  | ug/L   | 0.00     |
| Spiked Amount                 | 100.000 | Range | 40 - 130 | Recovery | =      | 125.130% |
| 24) Chloroform-d              | 5.049   | 84    | 112956   | 49.924   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =      | 99.840%  |
| 26) 1,2-Dichloroethane-d4     | 5.692   | 65    | 75840    | 54.808   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =      | 109.620% |
| 32) Benzene-d6                | 5.717   | 84    | 226283   | 53.291   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =      | 106.580% |
| 36) 1,2-Dichloropropane-d6    | 6.679   | 67    | 76310    | 54.411   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 120 | Recovery | =      | 108.820% |
| 41) Toluene-d8                | 7.891   | 98    | 184434   | 48.459   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =      | 96.920%  |
| 43) trans-1,3-Dichloroprop... | 8.174   | 79    | 31416    | 49.886   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =      | 99.780%  |
| 47) 2-Hexanone-d5             | 8.627   | 63    | 68525    | 118.137  | ug/L   | 0.00     |
| Spiked Amount                 | 100.000 | Range | 45 - 130 | Recovery | =      | 118.140% |
| 56) 1,1,2,2-Tetrachloroeth... | 10.749  | 84    | 120937   | 57.319   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 65 - 120 | Recovery | =      | 114.640% |
| 66) 1,2-Dichlorobenzene-d4    | 12.187  | 152   | 62615    | 55.255   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =      | 110.500% |
| Target Compounds              |         |       |          |          |        |          |
|                               |         |       |          |          |        | Qvalue   |
| 2) Dichlorodifluoromethane    | 1.377   | 85    | 2104     | 2.521    | ug/L   | 96       |
| 3) Chloromethane              | 1.515   | 50    | 3336     | 3.731    | ug/L   | 89       |
| 5) Vinyl chloride             | 1.599   | 62    | 2995     | 2.868    | ug/L # | 72       |
| 6) Bromomethane               | 1.824   | 94    | 2957     | 4.778    | ug/L   | 86       |
| 8) Chloroethane               | 1.898   | 64    | 2311     | 3.227    | ug/L   | 94       |
| 9) Trichlorofluoromethane     | 2.116   | 101   | 4149     | 2.922    | ug/L   | 99       |
| 10) 1,1,2-Trichloro-1,2,2-... | 2.563   | 101   | 3190     | 3.304    | ug/L # | 83       |
| 12) 1,1-Dichloroethene        | 2.563   | 96    | 2565     | 3.320    | ug/L # | 1        |
| 13) Acetone                   | 2.621   | 43    | 5654     | 7.077    | ug/L   | 98       |
| 14) Carbon disulfide          | 2.779   | 76    | 4730     | 3.254    | ug/L   | 97       |
| 15) Methyl Acetate            | 2.946   | 43    | 4850m    | 3.767    | ug/L   |          |
| 16) Methylene chloride        | 3.033   | 84    | 4641     | 3.766    | ug/L   | 96       |
| 17) trans-1,2-Dichloroethene  | 3.345   | 96    | 1836m    | 2.365    | ug/L   |          |
| 18) Methyl tert-butyl Ether   | 3.348   | 73    | 8935     | 2.908    | ug/L   | 97       |
| 19) 1,1-Dichloroethane        | 3.849   | 63    | 5892m    | 2.980    | ug/L   |          |
| 20) cis-1,2-Dichloroethene    | 4.656   | 96    | 2977     | 2.898    | ug/L   | 87       |
| 22) 2-Butanone                | 4.714   | 43    | 6892m    | 6.623    | ug/L   |          |
| 23) Bromochloromethane        | 4.968   | 128   | 1455m    | 2.717    | ug/L   |          |
| 27) 1,2-Dichloroethane        | 5.788   | 62    | 4968     | 3.118    | ug/L   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU052524\  
 Data File : VU058993.D  
 Acq On : 25 May 2024 12:54  
 Operator : MD/SY  
 Sample : P2409-05  
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA\_U/MEOH  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 MDL-MED-SOIL-QT2-2024-04

Quant Time: May 27 05:44:27 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM052324WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Mon May 27 04:10:55 2024  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Mahesh  
 Dadoda

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |    |
|-------------------------------|--------|------|----------|-------|-------|----------|----|
| 29) Cyclohexane               | 5.373  | 56   | 3005     | 2.835 | ug/L  | #        | 86 |
| 30) 1,1,1-Trichloroethane     | 5.309  | 97   | 4212     | 2.818 | ug/L  |          | 96 |
| 31) Carbon tetrachloride      | 5.512  | 117  | 2937m    | 2.432 | ug/L  |          |    |
| 34) Trichloroethene           | 6.537  | 95   | 2991     | 3.010 | ug/L  |          | 90 |
| 35) Methylcyclohexane         | 6.750  | 83   | 3293     | 2.827 | ug/L  |          | 93 |
| 37) 1,2-Dichloropropane       | 6.785  | 63   | 3686     | 3.029 | ug/L  | #        | 91 |
| 38) Bromodichloromethane      | 7.100  | 83   | 4158     | 2.749 | ug/L  | #        | 97 |
| 39) cis-1,3-Dichloropropene   | 7.602  | 75   | 4665     | 2.783 | ug/L  |          | 90 |
| 40) 4-Methyl-2-pentanone      | 7.788  | 43   | 11689    | 6.738 | ug/L  |          | 91 |
| 42) Toluene                   | 7.965  | 91   | 10603    | 2.642 | ug/L  |          | 99 |
| 44) trans-1,3-Dichloropropene | 8.206  | 75   | 4415     | 2.782 | ug/L  |          | 98 |
| 45) 1,1,2-Trichloroethane     | 8.396  | 97   | 3284     | 2.733 | ug/L  |          | 96 |
| 46) Tetrachloroethene         | 8.550  | 164  | 1564     | 2.360 | ug/L  |          | 94 |
| 48) 2-Hexanone                | 8.685  | 43   | 10687m   | 7.482 | ug/L  |          |    |
| 49) Dibromochloromethane      | 8.804  | 129  | 3126     | 2.723 | ug/L  |          | 78 |
| 50) 1,2-Dibromoethane         | 8.920  | 107  | 3331     | 2.902 | ug/L  | #        | 99 |
| 51) Chlorobenzene             | 9.441  | 112  | 7075     | 2.667 | ug/L  |          | 94 |
| 52) Ethylbenzene              | 9.566  | 91   | 10674    | 2.457 | ug/L  |          | 92 |
| 53) m,p-Xylene                | 9.692  | 106  | 3643     | 2.335 | ug/L  |          | 95 |
| 54) o-Xylene                  | 10.097 | 106  | 3725     | 2.385 | ug/L  |          | 77 |
| 55) Styrene                   | 10.116 | 104  | 5801     | 2.080 | ug/L  |          | 87 |
| 57) 1,1,2,2-Tetrachloroethane | 10.775 | 83   | 7553     | 3.489 | ug/L  |          | 93 |
| 59) Bromoform                 | 10.286 | 173  | 2065     | 2.837 | ug/L  | #        | 94 |
| 60) 1,2,3-Trichloropropane    | 10.817 | 75   | 5695     | 4.075 | ug/L  |          | 96 |
| 61) Isopropylbenzene          | 10.479 | 105  | 9214     | 2.627 | ug/L  |          | 97 |
| 62) 1,3,5-Trimethylbenzene    | 11.084 | 105  | 6810     | 2.465 | ug/L  |          | 97 |
| 63) 1,2,4-Trimethylbenzene    | 11.466 | 105  | 6365     | 2.284 | ug/L  |          | 96 |
| 64) 1,3-Dichlorobenzene       | 11.746 | 146  | 4780     | 2.625 | ug/L  |          | 86 |
| 65) 1,4-Dichlorobenzene       | 11.833 | 146  | 5342     | 2.905 | ug/L  |          | 98 |
| 67) 1,2-Dichlorobenzene       | 12.209 | 146  | 5731     | 3.095 | ug/L  |          | 94 |
| 68) 1,2-Dibromo-3-chloropr... | 12.994 | 75   | 1267     | 3.351 | ug/L  |          | 91 |
| 69) 1,3,5-Trichlorobenzene    | 13.219 | 180  | 3239     | 2.526 | ug/L  |          | 89 |
| 70) 1,2,4-trichlorobenzene    | 13.846 | 180  | 2154     | 1.950 | ug/L  | #        | 91 |
| 71) Naphthalene               | 14.100 | 128  | 7238m    | 2.247 | ug/L  |          |    |
| 72) 1,2,3-Trichlorobenzene    | 14.331 | 180  | 2520m    | 2.227 | ug/L  |          |    |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU052524\  
Data File : VU058993.D  
Acq On : 25 May 2024 12:54  
Operator : MD/SY  
Sample : P2409-05  
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA\_U/MEOH  
ALS Vial : 7 Sample Multiplier: 1


Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Quant Time: May 27 05:44:27 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM052324WMA.M  
Quant Title : VOC Analysis  
QLast Update : Mon May 27 04:10:55 2024  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By :Mahesh  
Dadoda





#2  
Dichlorodifluoromethane  
Concen: 2.521 ug/L  
RT: 1.377 min Scan# 2104  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 85 Resp: 2104  
Ion Ratio Lower Upper  
85 100  
87 29.8 25.7 38.5


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda




05/30/2024  
Supervised By :Semsettin  
Yesilyurt



05/30/2024




#3  
Chloromethane  
Concen: 3.731 ug/L  
RT: 1.515 min Scan# 70  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54




Tgt Ion: 50 Resp: 3336  
Ion Ratio Lower Upper  
50 100  
52 27.9 23.9 44.3





#5  
Vinyl chloride  
Concen: 2.868 ug/L  
RT: 1.599 min Scan# 90  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 62 Resp: 299  
Ion Ratio Lower Upper  
62 100  
64 49.7 23.7 43.9


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#6  
Bromomethane  
Concen: 4.778 ug/L  
RT: 1.824 min Scan# 166  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54


Tgt Ion: 94 Resp: 2957  
Ion Ratio Lower Upper  
94 100  
96 81.9 66.6 123.6





#8  
Chloroethane  
Concen: 3.227 ug/L  
RT: 1.898 min Scan# 189  
Delta R.T. -0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 64 Resp: 231  
Ion Ratio Lower Upper  
64 100  
66 28.9 22.7 42.1


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#9  
Trichlorofluoromethane  
Concen: 2.922 ug/L  
RT: 2.116 min Scan# 257  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:101 Resp: 4149  
Ion Ratio Lower Upper  
101 100  
103 64.0 51.6 77.4







#10  
1,1,2-Trichloro-1,2,2-trifluoroethane  
Concen: 3.304 ug/L  
RT: 2.563 min Scan# 397  
Delta R.T. -0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 101 Resp: 3190  
Ion Ratio Lower Upper  
101 100  
85 35.3 34.9 52.3  
151 52.2 54.5 81.7



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#12  
1,1-Dichloroethene  
Concen: 3.320 ug/L  
RT: 2.563 min Scan# 396  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 96 Resp: 2565  
Ion Ratio Lower Upper  
96 100  
61 376.9 122.2 227.0#  
63 1764.8 102.8 190.8#





#13

Acetone

Concen: 7.077 ug/L

RT: 2.621 min Scan# 41

Delta R.T. 0.006 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

Instrument :

MSVOA\_U



ClientSampleId :

MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 43 Resp: 5654  
Ion Ratio Lower Upper  
43 100  
58 34.7 0.0 67.2



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda



05/30/2024  
Supervised By :Semsettin  
Yesilyurt

05/30/2024



#14

Carbon disulfide

Concen: 3.254 ug/L




RT: 2.779 min Scan# 463

Delta R.T. 0.003 min


Lab File: VU058993.D

Acq: 25 May 2024 12:54

Tgt Ion: 76 Resp: 4730  
Ion Ratio Lower Upper  
76 100  
78 9.8 7.1 10.7









#15  
Methyl Acetate  
Concen: 3.767 ug/L m  
RT: 2.946 min Scan# 512  
Delta R.T. 0.009 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 43 Resp: 4850  
Ion Ratio Lower Upper  
43 100  
74 1.2 19.8 29.8



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#16  
Methylene chloride  
Concen: 3.766 ug/L  
RT: 3.033 min Scan# 542  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 84 Resp: 4641  
Ion Ratio Lower Upper  
84 100  
86 59.3 44.9 83.5  
49 124.6 85.4 158.6







#17  
trans-1,2-Dichloroethene  
Concen: 2.365 ug/L m  
RT: 3.345 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 96 Resp: 1830  
Ion Ratio Lower Upper  
96 100  
61 142.6 103.5 192.1  
98 61.6 44.9 83.3



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#18  
Methyl tert-butyl Ether  
Concen: 2.908 ug/L  
RT: 3.348 min Scan# 640  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 73 Resp: 8935  
Ion Ratio Lower Upper  
73 100  
43 20.9 15.4 23.2  
57 24.2 18.6 27.8







#19  
1,1-Dichloroethane  
Concen: 2.980 ug/L m  
RT: 3.849 min Scan# 798  
Delta R.T. -0.007 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 63 Resp: 5892  
Ion Ratio Lower Upper  
63 100  
65 28.7 22.3 41.3  
83 10.7 8.5 15.9



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#20  
cis-1,2-Dichloroethene  
Concen: 2.898 ug/L  
RT: 4.656 min Scan# 1047  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54


Tgt Ion: 96 Resp: 2977  
Ion Ratio Lower Upper  
96 100  
61 140.7 91.1 169.3  
98 45.7 44.4 82.4





#22  
2-Butanone  
Concen: 6.623 ug/L m  
RT: 4.714 min Scan# 1065  
Delta R.T. 0.026 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 43 Resp: 689  
Ion Ratio Lower Upper  
43 100  
72 6.9 13.9 41.7


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#23  
Bromochloromethane  
Concen: 2.717 ug/L m  
RT: 4.968 min Scan# 1144  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:128 Resp: 1455  
Ion Ratio Lower Upper  
128 100  
49 199.3 123.8 229.8  
130 132.2 90.6 168.2  
51 60.9 43.8 65.6







#27  
1,2-Dichloroethane  
Concen: 3.118 ug/L  
RT: 5.788 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 62 Resp: 4968  
Ion Ratio Lower Upper  
62 100  
98 9.8 7.0 10.4



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#29  
Cyclohexane  
Concen: 2.835 ug/L  
RT: 5.373 min Scan# 1270  
Delta R.T. -0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 56 Resp: 3005  
Ion Ratio Lower Upper  
56 100  
69 16.9 24.3 36.5#  
84 74.8 66.7 100.1







#30  
1,1,1-Trichloroethane  
Concen: 2.818 ug/L  
RT: 5.309 min Scan# 1248  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 97 Resp: 421  
Ion Ratio Lower Upper  
97 100  
99 61.5 51.8 77.8  
61 50.8 38.9 58.3



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#31  
Carbon tetrachloride  
Concen: 2.432 ug/L m  
RT: 5.512 min Scan# 1313  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54


Tgt Ion:117 Resp: 2937  
Ion Ratio Lower Upper  
117 100  
119 2.9 77.4 116.0#





#34  
Trichloroethene  
Concen: 3.010 ug/L  
RT: 6.537 min Scan# 1632  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 95 Resp: 299  
Ion Ratio Lower Upper  
95 100  
97 67.8 45.9 85.3  
132 97.6 60.5 112.5  
130 103.8 64.0 118.8


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#35  
Methylcyclohexane  
Concen: 2.827 ug/L  
RT: 6.750 min Scan# 1698  
Delta R.T. -0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 83 Resp: 3293  
Ion Ratio Lower Upper  
83 100  
55 74.0 61.5 92.3  
98 37.3 36.2 54.2







#37  
 1,2-Dichloropropane  
 Concen: 3.029 ug/L  
 RT: 6.785 min Scan# 1708  
 Delta R.T. 0.003 min  
 Lab File: VU058993.D  
 Acq: 25 May 2024 12:54

Instrument : MSVOA\_U  
 ClientSampleId : MDL-MED-SOIL-QT2-2024-04

|           |       |       |     |
|-----------|-------|-------|-----|
| Tgt Ion:  | 63    | Resp: | 368 |
| Ion Ratio | Lower | Upper |     |
| 63        | 100   |       |     |
| 112       | 0.0   | 2.5   | 3.7 |



Manual Integrations  
**APPROVED**

Reviewed By :Mahesh  
 Dadoda






05/30/2024  
 Supervised By :Semsettin  
 Yesilyurt

05/30/2024




#38  
 Bromodichloromethane  
 Concen: 2.749 ug/L  
 RT: 7.100 min Scan# 1807  
 Delta R.T. 0.003 min  
 Lab File: VU058993.D  
 Acq: 25 May 2024 12:54

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 83    | Resp: | 4158 |
| Ion Ratio | Lower | Upper |      |
| 83        | 100   |       |      |
| 85        | 60.3  | 43.5  | 80.9 |
| 127       | 9.3   | 5.8   | 8.6  |









#39  
cis-1,3-Dichloropropene  
Concen: 2.783 ug/L  
RT: 7.602 min Scan# 1963  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 75 Resp: 466  
Ion Ratio Lower Upper  
75 100  
77 37.7 22.5 41.7



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#40  
4-Methyl-2-pentanone  
Concen: 6.738 ug/L  
RT: 7.788 min Scan# 2021  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion: 43 Resp: 11689  
Ion Ratio Lower Upper  
43 100  
58 35.9 33.5 50.3  
100 12.0 11.8 17.8





#42

Toluene

Concen: 2.642 ug/L

RT: 7.965 min Scan# 2075

Delta R.T. 0.003 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

Instrument :

MSVOA\_U

ClientSampleId :

MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 91 Resp: 1060



Ion Ratio Lower Upper

91 100

92 54.5 38.7 71.9

Manual Integrations



APPROVED

Reviewed By :Mahesh  
Dadoda

05/30/2024

Supervised By :Semsettin  
Yesilyurt

05/30/2024



#44

trans-1,3-Dichloropropene

Concen: 2.782 ug/L

RT: 8.206 min Scan# 2151

Delta R.T. 0.003 min

Lab File: VU058993.D




Acq: 25 May 2024 12:54


Tgt Ion: 75 Resp: 4415

Ion Ratio Lower Upper

75 100

77 31.9 23.0 42.8





#45

1,1,2-Trichloroethane

Concen: 2.733 ug/L

RT: 8.396 min Scan# 21

Delta R.T. 0.003 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

Instrument :

MSVOA\_U

ClientSampleId :



MDL-MED-SOIL-QT2-2024-04

| Tgt Ion: | 97    | Resp: | 3284  |
|----------|-------|-------|-------|
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 99       | 60.1  | 42.6  | 79.2  |
| 83       | 83.8  | 62.2  | 115.6 |
| 85       | 61.7  | 40.2  | 74.6  |

Manual Integrations



APPROVED

Reviewed By :Mahesh  
Dadoda



05/30/2024  
Supervised By :Semsettin  
Yesilyurt

05/30/2024



#46

Tetrachloroethene

Concen: 2.360 ug/L




RT: 8.550 min Scan# 2258


Delta R.T. 0.003 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

| Tgt Ion: | 164   | Resp: | 1564  |
|----------|-------|-------|-------|
| Ion      | Ratio | Lower | Upper |
| 164      | 100   |       |       |
| 129      | 96.8  | 70.6  | 131.0 |
| 131      | 111.1 | 69.2  | 128.6 |
| 166      | 133.9 | 91.9  | 170.7 |







#48  
2-Hexanone  
Concen: 7.482 ug/L m  
RT: 8.685 min Scan# 2106  
Delta R.T. 0.009 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

|             |             |
|-------------|-------------|
| Tgt Ion: 43 | Resp: 1068  |
| Ion Ratio   | Lower Upper |
| 43 100      |             |
| 58 62.0     | 47.8 71.6   |
| 57 19.6     | 15.7 23.5   |
| 100 10.7    | 9.0 13.4    |



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#49  
Dibromochloromethane  
Concen: 2.723 ug/L  
RT: 8.804 min Scan# 2337  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54


|             |             |
|-------------|-------------|
| Tgt Ion:129 | Resp: 3126  |
| Ion Ratio   | Lower Upper |
| 129 100     |             |
| 127 61.4    | 56.8 105.4  |





#50  
1,2-Dibromoethane  
Concen: 2.902 ug/L  
RT: 8.920 min Scan# 21  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04



Tgt Ion:107 Resp: 333  
Ion Ratio Lower Upper  
107 100  
109 97.8 68.0 126.2  
188 0.0 2.6 3.8


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#51  
Chlorobenzene  
Concen: 2.667 ug/L  
RT: 9.441 min Scan# 2535  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:112 Resp: 7075  
Ion Ratio Lower Upper  
112 100  
114 34.1 21.8 40.6  
77 66.3 48.8 73.2







#52  
Ethylbenzene  
Concen: 2.457 ug/L  
RT: 9.566 min Scan# 2106  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion: 91 Resp: 10674  
Ion Ratio Lower Upper  
91 100  
106 33.8 20.8 38.6



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#53  
m,p-Xylene  
Concen: 2.335 ug/L  
RT: 9.692 min Scan# 2613  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:106 Resp: 3643  
Ion Ratio Lower Upper  
106 100  
91 216.2 145.7 270.7







#54  
o-Xylene  
Concen: 2.385 ug/L  
RT: 10.097 min Scan# 2738  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04

Tgt Ion:106 Resp: 3725  
Ion Ratio Lower Upper  
106 100  
91 256.7 154.0 286.0



Manual Integrations  
**APPROVED**

Reviewed By :Mahesh  
Dadoda






05/30/2024  
Supervised By :Semsettin  
Yesilyurt


05/30/2024



#55  
Styrene  
Concen: 2.080 ug/L  
RT: 10.116 min Scan# 2745  
Delta R.T. 0.009 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:104 Resp: 5801  
Ion Ratio Lower Upper  
104 100  
78 54.4 32.3 59.9








#57  
 1,1,2,2-Tetrachloroethane  
 Concen: 3.489 ug/L  
 RT: 10.775 min Scan# 2950  
 Delta R.T. -0.000 min  
 Lab File: VU058993.D  
 Acq: 25 May 2024 12:54

Instrument : MSVOA\_U  
 ClientSampleId : MDL-MED-SOIL-QT2-2024-04

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 83    | Resp: | 755  |
| Ion Ratio | Lower | Upper |      |
| 83        | 100   |       |      |
| 85        | 68.5  | 43.5  | 80.9 |
| 131       | 9.9   | 7.1   | 10.7 |


Manual Integrations  
**APPROVED**

Reviewed By :Mahesh Dadoda






05/30/2024  
 Supervised By :Semsettin Yesilyurt

05/30/2024




#59  
 Bromoform  
 Concen: 2.837 ug/L  
 RT: 10.286 min Scan# 2798  
 Delta R.T. 0.003 min  
 Lab File: VU058993.D  
 Acq: 25 May 2024 12:54

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 173   | Resp: | 2065 |
| Ion Ratio | Lower | Upper |      |
| 173       | 100   |       |      |
| 175       | 46.3  | 40.6  | 61.0 |
| 254       | 6.8   | 7.3   | 10.9 |








#60  
1,2,3-Trichloropropane  
Concen: 4.075 ug/L  
RT: 10.817 min Scan# 2963  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04



Tgt Ion: 75 Resp: 569  
Ion Ratio Lower Upper  
75 100  
77 32.0 25.8 38.6  
110 31.2 28.3 42.5


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda




05/30/2024  
Supervised By :Semsettin  
Yesilyurt



05/30/2024




#61  
Isopropylbenzene  
Concen: 2.627 ug/L  
RT: 10.479 min Scan# 2858  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54




Tgt Ion:105 Resp: 9214  
Ion Ratio Lower Upper  
105 100  
120 23.3 20.0 30.0  
77 17.6 13.4 20.0





#62  
1,3,5-Trimethylbenzene  
Concen: 2.465 ug/L  
RT: 11.084 min Scan# 3046  
Delta R.T. -0.000 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04



Tgt Ion:105 Resp: 6810  
Ion Ratio Lower Upper  
105 100  
120 48.4 36.9 55.3


Manual Integrations  
**APPROVED**

Reviewed By :Mahesh  
Dadoda




05/30/2024  
Supervised By :Semsettin  
Yesilyurt



05/30/2024




#63  
1,2,4-Trimethylbenzene  
Concen: 2.284 ug/L  
RT: 11.466 min Scan# 3165  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54




Tgt Ion:105 Resp: 6365  
Ion Ratio Lower Upper  
105 100  
120 47.0 35.3 52.9





#64  
1,3-Dichlorobenzene  
Concen: 2.625 ug/L  
RT: 11.746 min Scan# 31  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL-MED-SOIL-QT2-2024-04



Tgt Ion:146 Resp: 4780  
Ion Ratio Lower Upper  
146 100  
111 46.0 28.6 53.0  
148 76.3 44.2 82.2


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda




05/30/2024  
Supervised By :Semsettin  
Yesilyurt



05/30/2024




#65  
1,4-Dichlorobenzene  
Concen: 2.905 ug/L  
RT: 11.833 min Scan# 3279  
Delta R.T. 0.003 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Tgt Ion:146 Resp: 5342  
Ion Ratio Lower Upper  
146 100  
111 41.6 27.3 50.7  
148 64.4 44.7 83.1





#67

1,2-Dichlorobenzene

Concen: 3.095 ug/L

RT: 12.209 min Scan# 31

Delta R.T. 0.003 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

Instrument :

MSVOA\_U



ClientSampleId :

MDL-MED-SOIL-QT2-2024-04

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 146   | Resp: | 573   |
| Ion      | Ratio | Lower | Upper |
| 146      | 100   |       |       |
| 111      | 48.5  | 34.1  | 51.1  |
| 148      | 61.9  | 52.2  | 78.4  |



Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda



05/30/2024  
Supervised By :Semsettin  
Yesilyurt

05/30/2024



#68

1,2-Dibromo-3-chloropropane

Concen: 3.351 ug/L




RT: 12.994 min Scan# 3640


Delta R.T. 0.003 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 75    | Resp: | 1267  |
| Ion      | Ratio | Lower | Upper |
| 75       | 100   |       |       |
| 155      | 64.6  | 60.2  | 90.4  |
| 157      | 89.8  | 77.2  | 115.8 |





#69

1,3,5-Trichlorobenzene

Concen: 2.526 ug/L

RT: 13.219 min Scan# 31

Delta R.T. 0.006 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54

Instrument :

MSVOA\_U

ClientSampleId :

MDL-MED-SOIL-QT2-2024-04

Tgt Ion:180 Resp: 3239

Ion Ratio Lower Upper

180 100

182 80.5 75.7 113.5

184 26.8 24.3 36.5



145 26.7 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :Mahesh

Dadoda





05/30/2024

Supervised By :Semsettin

Yesilyurt

05/30/2024



#70

1,2,4-trichlorobenzene

Concen: 1.950 ug/L

RT: 13.846 min Scan# 3905

Delta R.T. 0.009 min

Lab File: VU058993.D

Acq: 25 May 2024 12:54




Tgt Ion:180 Resp: 2154


Ion Ratio Lower Upper

180 100

182 101.2 78.6 117.8


145 15.5 24.6 37.0#





#71  
Naphthalene  
Concen: 2.247 ug/L m  
RT: 14.100 min Scan# 3978  
Delta R.T. 0.019 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54



Instrument : MSVOA\_U  
ClientSampleId : MDL-MED-SOIL-QT2-2024-04



Tgt Ion:128 Resp: 7238  
Ion Ratio Lower Upper  
128 100  
127 5.8 10.8 16.2  
129 2.8 8.8 13.2


Manual Integrations  
APPROVED

Reviewed By :Mahesh  
Dadoda



05/30/2024  
Supervised By :Semsettin  
Yesilyurt

05/30/2024



#72  
1,2,3-Trichlorobenzene  
Concen: 2.227 ug/L m  
RT: 14.331 min Scan# 4056  
Delta R.T. 0.006 min  
Lab File: VU058993.D  
Acq: 25 May 2024 12:54

Tgt Ion:180 Resp: 2520  
Ion Ratio Lower Upper  
180 100  
182 105.4 76.1 114.1  
145 22.9 26.2 39.2

