

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052524\
 Data File : VU059012.D
 Acq On : 25 May 2024 20:40
 Operator : MD/SY
 Sample : P2568-21ME
 Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH642ME

Quant Time: May 27 04:17:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon May 27 04:10:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	145801	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	144172	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	59594	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	54785	35.527	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.060%
7) Chloroethane-d5	1.872	69	48633	37.948	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.900%
11) 1,1-Dichloroethene-d2	2.534	63	72598	30.815	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.620%
21) 2-Butanone-d5	4.624	46	118295	111.679	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.680%
24) Chloroform-d	5.049	84	113391	42.870	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.740%
26) 1,2-Dichloroethane-d4	5.692	65	78478	48.514	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.020%
32) Benzene-d6	5.714	84	229327	46.657	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
36) 1,2-Dichloropropane-d6	6.682	67	80720	49.721	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.440%
41) Toluene-d8	7.891	98	180477	40.965	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.920%
43) trans-1,3-Dichloroprop...	8.174	79	30380	41.675	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.360%
47) 2-Hexanone-d5	8.643	63	72038	107.290	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.290%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	120778	49.452	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.900%
66) 1,2-Dichlorobenzene-d4	12.190	152	58156	47.295	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052524\
Data File : VU059012.D
Acq On : 25 May 2024 20:40
Operator : MD/SY
Sample : P2568-21ME
Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH642ME

Quant Time: May 27 04:17:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
Quant Title : VOC Analysis
QLast Update : Mon May 27 04:10:55 2024
Response via : Initial Calibration

