

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU060220\
 Data File : VU038493.D
 Acq On : 02 Jun 2020 15:09
 Operator : JC/MD
 Sample : L2844-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF1

Quant Time: Jun 03 01:32:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 03 00:48:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	321285	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	322593	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	149160	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	65204	3.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.80%
7) Chloroethane-d5	1.93	69	63649	3.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.60%
11) 1,1-Dichloroethene-d2	2.59	63	108962	2.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.20%#
20) 2-Butanone-d5	4.68	46	339115	53.68	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.36%
24) Chloroform-d	5.10	84	173567	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
26) 1,2-Dichloroethane-d4	5.74	65	107441	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.76	84	330038	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	6.72	67	105786	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.92	98	268159	3.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.20%
43) trans-1,3-Dichloropropene-	8.20	79	43395	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.66	63	289954	55.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.26%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	101505	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	12.20	152	119277	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%

Target Compounds					Qvalue
13) Acetone	2.69	43	15720	3.80 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU060220\
Data File : VU038493.D
Acq On : 02 Jun 2020 15:09
Operator : JC/MD
Sample : L2844-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF1

Quant Time: Jun 03 01:32:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 03 00:48:29 2020
Response via : Initial Calibration

