

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060324\
 Data File : VU059128.D
 Acq On : 03 Jun 2024 17:32
 Operator : MD/SY
 Sample : P2659-18ME
 Misc : 4.87g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHC39ME

Quant Time: Jun 04 07:00:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 01 01:17:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	218385	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	227334	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	105768	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	81235	62.127	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	124.260%
7) Chloroethane-d5	1.872	69	69582	62.122	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	124.240%
11) 1,1-Dichloroethene-d2	2.538	63	117925	45.690	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.380%
21) 2-Butanone-d5	4.618	46	232651	188.769	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	188.770%#
24) Chloroform-d	5.052	84	194488	74.717	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	149.440%#
26) 1,2-Dichloroethane-d4	5.689	65	129512	81.178	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	162.360%#
32) Benzene-d6	5.714	84	375072	71.885	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	143.760%#
36) 1,2-Dichloropropane-d6	6.682	67	138332	76.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	153.780%#
41) Toluene-d8	7.891	98	312377	70.525	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	141.040%#
43) trans-1,3-Dichloroprop...	8.174	79	55306	76.053	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	152.100%#
47) 2-Hexanone-d5	8.637	63	150193	180.536	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	180.540%#
56) 1,1,2,2-Tetrachloroeth...	10.753	84	243201	79.876	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	159.760%#
66) 1,2-Dichlorobenzene-d4	12.190	152	128618	81.486	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	162.980%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060324\
Data File : VU059128.D
Acq On : 03 Jun 2024 17:32
Operator : MD/SY
Sample : P2659-18ME
Misc : 4.87g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHC39ME

Quant Time: Jun 04 07:00:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 01 01:17:21 2024
Response via : Initial Calibration

