

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060620\
 Data File : VU038622.D
 Acq On : 06 Jun 2020 17:16
 Operator : SY/MD
 Sample : L2910-15
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D36

Quant Time: Jun 07 05:50:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jun 07 04:01:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	303286	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	303571	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	160570	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	66384	3.38	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.60%
7) Chloroethane-d5	1.93	69	63008	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.59	63	112499	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	4.67	46	342634	57.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.92%
24) Chloroform-d	5.10	84	180102	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.74	65	110142	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.76	84	348062	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.72	67	113778	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.92	98	292447	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	8.20	79	43508	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.65	63	295057	59.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.22%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	106296	5.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.40%
64) 1,2-Dichlorobenzene-d4	12.20	152	132798	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%

Target Compounds					Ovalue
13) Acetone	2.67	43	21598	5.535 ug/L	97
27) 1,2-Dichloroethane	5.83	62	31567	1.225 ug/L	98
30) Cyclohexane	5.42	56	10900	0.324 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060620\
Data File : VU038622.D
Acq On : 06 Jun 2020 17:16
Operator : SY/MD
Sample : L2910-15
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D36

Quant Time: Jun 07 05:50:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Jun 07 04:01:21 2020
Response via : Initial Calibration

