

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060620\
 Data File : VU038624.D
 Acq On : 06 Jun 2020 18:02
 Operator : SY/MD
 Sample : L2910-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D45

Quant Time: Jun 07 05:59:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jun 07 04:01:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	311969	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	302176	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	147536	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	60948	3.02	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.40%
7) Chloroethane-d5	1.93	69	60190	3.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.60%
11) 1,1-Dichloroethene-d2	2.59	63	105907	2.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.40%#
20) 2-Butanone-d5	4.67	46	332363	54.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.38%
24) Chloroform-d	5.10	84	172776	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
26) 1,2-Dichloroethane-d4	5.74	65	103724	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.76	84	341628	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.72	67	108697	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.92	98	277455	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
43) trans-1,3-Dichloropropene-	8.20	79	41045	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
46) 2-Hexanone-d5	8.65	63	287188	58.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.58%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	102279	5.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.80%
64) 1,2-Dichlorobenzene-d4	12.20	152	126742	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds					Ovalue
13) Acetone	2.67	43	35077	8.739 ug/L	97
30) Cyclohexane	5.42	56	9008	0.269 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU060620\
Data File : VU038624.D
Acq On : 06 Jun 2020 18:02
Operator : SY/MD
Sample : L2910-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D45

Quant Time: Jun 07 05:59:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR053120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Jun 07 04:01:21 2020
Response via : Initial Calibration

