

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060620\  
 Data File : VU038610.D  
 Acq On : 06 Jun 2020 12:18  
 Operator : SY/MD  
 Sample : VU0606WBL01  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK05

Quant Time: Jun 07 04:26:10 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR053120WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Sun Jun 07 04:01:21 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 6.28  | 114  | 321679   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.44  | 117  | 315419   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.82 | 152  | 149021   | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.61   | 65    | 92642    | 4.45     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 89.00%  |
| 7) Chloroethane-d5             | 1.93   | 69    | 77914    | 4.87     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 97.40%  |
| 11) 1,1-Dichloroethene-d2      | 2.59   | 63    | 145022   | 3.54     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 70.80%  |
| 20) 2-Butanone-d5              | 4.67   | 46    | 338790   | 53.57    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 107.14% |
| 24) Chloroform-d               | 5.10   | 84    | 192869   | 4.93     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 98.60%  |
| 26) 1,2-Dichloroethane-d4      | 5.74   | 65    | 111969   | 5.04     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 100.80% |
| 32) Benzene-d6                 | 5.76   | 84    | 392383   | 4.98     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 99.60%  |
| 36) 1,2-Dichloropropane-d6     | 6.72   | 67    | 119868   | 4.95     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 99.00%  |
| 41) Toluene-d8                 | 7.92   | 98    | 324409   | 4.59     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 91.80%  |
| 43) trans-1,3-Dichloropropene- | 8.20   | 79    | 47981    | 4.61     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 92.20%  |
| 46) 2-Hexanone-d5              | 8.65   | 63    | 292610   | 56.90    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 113.80% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.77  | 84    | 106088   | 5.16     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 103.20% |
| 64) 1,2-Dichlorobenzene-d4     | 12.20  | 152   | 129989   | 5.06     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 101.20% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU060620\  
Data File : VU038610.D  
Acq On : 06 Jun 2020 12:18  
Operator : SY/MD  
Sample : VU0606WBL01  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VBLK05

Quant Time: Jun 07 04:26:10 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR053120WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Sun Jun 07 04:01:21 2020  
Response via : Initial Calibration

