

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060624\
 Data File : VU059162.D
 Acq On : 06 Jun 2024 12:46
 Operator : MD/SY
 Sample : VU0606MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK067

Quant Time: Jun 07 05:08:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 07 05:05:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	193934	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	192455	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	81080	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	50208	43.239	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.480%
7) Chloroethane-d5	1.891	69	46587	46.836	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.680%
11) 1,1-Dichloroethene-d2	2.554	63	67195	29.317	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.640%#
21) 2-Butanone-d5	4.608	46	117433	107.296	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.300%
24) Chloroform-d	5.052	84	113620	49.153	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.300%
26) 1,2-Dichloroethane-d4	5.692	65	72752	51.350	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.700%
32) Benzene-d6	5.718	84	210878	47.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.480%
36) 1,2-Dichloropropane-d6	6.682	67	76502	50.231	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.460%
41) Toluene-d8	7.891	98	168780	45.011	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.020%
43) trans-1,3-Dichloroprop...	8.174	79	27596	44.826	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.660%
47) 2-Hexanone-d5	8.624	63	63625	90.339	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.340%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	126695	49.152	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.300%
66) 1,2-Dichlorobenzene-d4	12.187	152	62709	51.826	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU060624\
Data File : VU059162.D
Acq On : 06 Jun 2024 12:46
Operator : MD/SY
Sample : VU0606MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK067

Quant Time: Jun 07 05:08:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 07 05:05:14 2024
Response via : Initial Calibration

