

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061024\
 Data File : VU059201.D
 Acq On : 10 Jun 2024 11:59
 Operator : MD/SY
 Sample : P2754-11ME
 Misc : 4.63g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4BQ0ME

Quant Time: Jun 11 06:02:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 11 05:56:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	302921	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	297022	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	137438	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	49104	27.074	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	54.140%#
7) Chloroethane-d5	1.911	69	49903	32.120	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	64.240%#
11) 1,1-Dichloroethene-d2	2.563	63	75264	21.023	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	42.040%#
21) 2-Butanone-d5	4.611	46	132161	77.308	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.310%
24) Chloroform-d	5.055	84	142576	39.488	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.980%
26) 1,2-Dichloroethane-d4	5.695	65	89941	40.642	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.280%
32) Benzene-d6	5.721	84	264434	38.790	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.580%
36) 1,2-Dichloropropane-d6	6.682	67	93782	39.898	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.800%
41) Toluene-d8	7.891	98	219225	37.882	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.760%#
43) trans-1,3-Dichloroprop...	8.174	79	37161	39.112	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.220%
47) 2-Hexanone-d5	8.624	63	80750	74.290	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.290%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	164814	41.430	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.860%
66) 1,2-Dichlorobenzene-d4	12.187	152	94965	46.301	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061024\
Data File : VU059201.D
Acq On : 10 Jun 2024 11:59
Operator : MD/SY
Sample : P2754-11ME
Misc : 4.63g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4BQ0ME

Quant Time: Jun 11 06:02:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 11 05:56:14 2024
Response via : Initial Calibration

