



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061324\
Data File : VU059224.D
Acq On : 13 Jun 2024 08:13
Operator : MD/SY
Sample : BFB161
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB161

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051424WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Jun 14 06:53:55 2024



AutoFind: Scans 2904, 2905, 2906; Background Corrected with Scan 2890

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	26728	PASS
75	95	30	80	50.8	74032	PASS
95	95	100	100	100.0	145600	PASS
96	95	5	9	6.3	9225	PASS
173	174	0.00	2	1.4	1685	PASS
174	95	50	120	81.3	118411	PASS
175	174	5	9	7.1	8463	PASS
176	174	95	101	96.9	114760	PASS
177	176	5	9	6.2	7098	PASS

m/z	Abundance
39.70	150.0
44.00	269.0

Instrument :
MSVOA_U
ClientSampleId :
BFB161

m/z	Abundance
35.80	1234.0
36.90	6160.0
37.90	6100.0
39.00	2340.0
39.80	183.0
44.00	1086.0
44.90	1193.0
46.90	1393.0
47.90	1027.0
49.00	5928.0
49.90	27824.0
50.90	8861.0
51.90	441.0
53.00	112.0
54.80	442.0
55.90	2290.0
56.90	3974.0
57.90	171.0
59.90	1447.0
60.90	7743.0
61.90	7233.0
63.00	6354.0
63.90	1093.0
64.90	1282.0
65.90	209.0
67.00	411.0
67.90	15200.0
69.00	15815.0
69.90	1308.0
71.90	853.0
72.90	6722.0
73.90	25312.0
75.00	76320.0
76.00	6688.0
76.90	1092.0
77.90	705.0
78.80	5673.0
79.90	1705.0
80.90	5774.0
81.80	1238.0
82.90	168.0
86.00	204.0
86.90	7554.0
87.90	6416.0
90.80	600.0
91.90	4249.0
92.90	6936.0
94.00	18384.0
95.00	149760.0
96.00	9410.0
96.90	338.0
103.90	698.0
104.70	186.0
105.80	610.0
106.80	170.0
110.90	135.0
111.70	126.0
112.90	189.0
114.80	172.0
115.80	509.0
116.90	1077.0
117.80	512.0
118.90	908.0
127.80	610.0
128.80	187.0
129.80	455.0
130.90	227.0
134.90	245.0
136.80	212.0
139.90	143.0
140.90	1576.0
141.90	190.0
142.90	2061.0
143.80	174.0
144.90	590.0
145.80	178.0
147.80	442.0
149.80	202.0
152.80	110.0
154.80	316.0
156.90	304.0
158.80	135.0
160.80	121.0
171.10	248.0
171.90	1529.0
173.00	1749.0
173.90	119984.0
174.90	8410.0
175.90	114736.0
176.90	7450.0
177.90	147.0

Instrument :
MSVOA_U
ClientSampleId :
BFB161

m/z	Abundance
35.80	1193.0
36.90	6094.0
37.90	6189.0
38.90	2351.0
40.10	175.0
43.90	1220.0
44.90	1239.0
46.90	1451.0
47.90	1021.0
48.90	5761.0
49.90	27648.0
51.00	9042.0
52.00	413.0
53.00	123.0
54.80	377.0
55.90	2375.0
56.90	3966.0
57.90	177.0
59.90	1534.0
60.90	7571.0
61.90	7253.0
63.00	6056.0
63.90	1080.0
64.90	1368.0
65.90	210.0
67.00	384.0
67.90	15156.0
69.00	15718.0
69.90	1162.0
71.90	771.0
72.90	6500.0
73.90	24848.0
75.00	76616.0
76.00	6534.0
76.90	1030.0
77.90	775.0
78.80	5421.0
79.90	1748.0
80.90	5846.0
81.90	1357.0
83.00	163.0
85.50	136.0
86.00	154.0
86.90	7505.0
87.90	6491.0
90.80	563.0
91.90	4166.0
92.90	7086.0
94.00	18552.0
95.00	150656.0
96.00	9540.0
96.90	277.0
102.90	120.0
103.90	650.0
104.80	206.0
105.80	625.0
106.80	196.0
110.90	150.0
111.70	138.0
112.90	209.0
114.90	167.0
115.90	549.0
116.90	1053.0
117.80	526.0
118.90	957.0
127.80	579.0
128.80	198.0
129.80	435.0
130.90	194.0
134.90	275.0
136.80	222.0
139.90	128.0
140.90	1553.0
141.80	176.0
142.90	2088.0
143.80	154.0
144.80	511.0
145.90	182.0
147.80	515.0
149.80	179.0
152.90	105.0
154.80	325.0
156.90	282.0
158.80	162.0
160.70	138.0
170.70	235.0
171.90	1570.0
173.00	1747.0
173.90	122056.0
174.90	8728.0
175.90	118712.0
176.90	7237.0

Instrument :
MSVOA_U
ClientSampleId :
BFB161

177.90

195.0

Instrument :
MSVOA_U
ClientSampleId :
BFB161

m/z	Abundance
35.90	1000.0
36.90	5324.0
37.90	5414.0
38.90	2192.0
40.10	185.0
43.90	1274.0
44.90	1218.0
46.90	1301.0
47.90	900.0
48.90	5183.0
49.90	24712.0
51.00	8329.0
51.90	360.0
53.00	127.0
54.90	344.0
55.90	2148.0
56.90	3535.0
57.90	163.0
59.90	1416.0
60.90	6640.0
61.90	6717.0
62.90	5184.0
63.90	912.0
64.90	1427.0
65.90	146.0
66.90	366.0
67.90	13634.0
68.90	13751.0
69.90	950.0
71.90	607.0
72.90	5706.0
73.90	22424.0
75.00	69160.0
76.00	5734.0
76.90	846.0
77.90	670.0
78.80	4735.0
79.90	1515.0
80.90	5181.0
81.90	1281.0
83.10	146.0
85.50	110.0
86.00	161.0
86.90	6672.0
87.90	6101.0
90.80	417.0
91.90	3809.0
92.90	6307.0
94.00	16832.0
95.00	136384.0
96.00	8725.0
96.90	228.0
103.80	554.0
104.80	218.0
105.90	578.0
106.90	193.0
110.80	144.0
111.70	139.0
112.90	179.0
114.90	139.0
115.90	551.0
116.90	1013.0
117.80	476.0
118.90	942.0
127.80	486.0
128.80	187.0
129.80	411.0
130.80	157.0
134.90	267.0
136.70	206.0
140.90	1361.0
141.70	175.0
142.90	1926.0
143.70	113.0
144.80	382.0
145.90	166.0
147.80	457.0
149.80	138.0
154.90	311.0
156.90	266.0
158.80	206.0
160.70	137.0
170.70	242.0
171.90	1468.0
173.00	1558.0
173.90	113192.0
174.90	8251.0
175.90	110832.0
176.90	6607.0
177.90	227.0

Instrument :
MSVOA_U
ClientSampleId :
BFB161