

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024530.D
 Acq On : 14 Jun 2018 11:29
 Operator : MD/SY
 Sample : J3456-01 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ES-OILY-WATER

Quant Time: Jun 15 02:50:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	174993	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	293716	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	279285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	161513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	129286	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
35) Dibromofluoromethane	4.89	113	118059	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
50) Toluene-d8	7.57	98	432187	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
62) 4-Bromofluorobenzene	10.31	95	165032	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
Target Compounds						
52) Toluene	7.64	92	2215	0.350	ug/l	90
68) m/p-Xylenes	9.38	106	1600	0.333	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024530.D
Acq On : 14 Jun 2018 11:29
Operator : MD/SY
Sample : J3456-01 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ES-OILY-WATER

Quant Time: Jun 15 02:50:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

