

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024531.D
 Acq On : 14 Jun 2018 11:52
 Operator : MD/SY
 Sample : J3246-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OK-02-061218

Quant Time: Jun 15 02:32:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	178480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	309001	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	294258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	172633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	135769	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
35) Dibromofluoromethane	4.89	113	122572	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	7.57	98	447194	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
62) 4-Bromofluorobenzene	10.31	95	175266	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
Target Compounds						
16) Acetone	2.32	43	18255	11.661	ug/l	97
18) Methyl Acetate	2.62	43	14088	4.198	ug/l	91
20) Methylene Chloride	2.71	84	37970	15.457	ug/l	96
84) 1,2,4-Trimethylbenzene	11.15	105	14342	1.225	ug/l	100
95) Naphthalene	13.75	128	193948	15.751	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024531.D
Acq On : 14 Jun 2018 11:52
Operator : MD/SY
Sample : J3246-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OK-02-061218

Quant Time: Jun 15 02:32:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

