

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024534.D
 Acq On : 14 Jun 2018 13:02
 Operator : MD/SY
 Sample : J3476-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 Z5TB06-180610

Quant Time: Jun 15 02:39:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	177310	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	300807	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	289726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	171887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	135861	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
35) Dibromofluoromethane	4.89	113	121406	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
50) Toluene-d8	7.57	98	444359	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	10.31	95	176399	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
Target Compounds						
16) Acetone	2.33	43	6943	4.464	ug/l	96
20) Methylene Chloride	2.71	84	1323	0.542	ug/l #	77
95) Naphthalene	13.76	128	4748	1.050	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024534.D
Acq On : 14 Jun 2018 13:02
Operator : MD/SY
Sample : J3476-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
Z5TB06-180610

Quant Time: Jun 15 02:39:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

