

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024540.D
 Acq On : 14 Jun 2018 15:21
 Operator : MD/SY
 Sample : J3443-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T3-MW-2-9.0-061118

Manual Integrations
 APPROVED

apatel
 6/15/2018 3:19:15 PM

Quant Time: Jun 15 03:47:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	316030	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	303783	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	183565	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	140915	45.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.14%	
35) Dibromofluoromethane	4.89	113	128793	49.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.22%	
50) Toluene-d8	7.57	98	472019	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
62) 4-Bromofluorobenzene	10.31	95	179750	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	
Target Compounds						
					Qvalue	
16) Acetone	2.32	43	14871	8.950	ug/l	95
31) Cyclohexane	5.00	56	40249	8.277	ug/l	96
39) Methylcyclohexane	6.42	83	109230	23.570	ug/l	98
40) Benzene	5.40	78	3105	0.293	ug/l	96
52) Toluene	7.64	92	4542	0.667	ug/l	96
67) Ethyl Benzene	9.26	91	26781	1.968	ug/l	96
73) Isopropylbenzene	10.17	105	506901	35.967	ug/l	99
78) n-propylbenzene	10.59	91	2136814	128.709	ug/l	98
80) 1,3,5-Trimethylbenzene	10.78	105	26805m	2.194	ug/l	
83) tert-Butylbenzene	11.10	119	25426	2.164	ug/l	91
84) 1,2,4-Trimethylbenzene	11.15	105	54047	4.342	ug/l	100
85) sec-Butylbenzene	11.33	105	177191	12.047	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024540.D
Acq On : 14 Jun 2018 15:21
Operator : MD/SY
Sample : J3443-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T3-MW-2-9.0-061118

Quant Time: Jun 15 03:47:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

apatel
6/15/2018 3:19:15 PM

