

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024551.D
 Acq On : 14 Jun 2018 19:35
 Operator : MD/SY
 Sample : J3448-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 N-48969

Quant Time: Jun 15 05:00:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	183654	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	311419	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	295780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	174067	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.32	65	136578	45.55	ug/l	0.00
Spiked Amount			Recovery	=	91.10%	
35) Dibromofluoromethane	4.89	113	123888	47.93	ug/l	0.00
Spiked Amount			Recovery	=	95.86%	
50) Toluene-d8	7.57	98	455783	48.35	ug/l	0.00
Spiked Amount			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	10.31	95	173398	46.33	ug/l	0.00
Spiked Amount			Recovery	=	92.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	33291	20.666	ug/l	99
25) 2-Butanone	4.28	43	4867	2.249	ug/l	92
29) Tetrahydrofuran	4.65	42	3786	2.825	ug/l #	89
30) Chloroform	4.68	83	3241	0.723	ug/l	90
52) Toluene	7.64	92	11710	1.746	ug/l	97
67) Ethyl Benzene	9.26	91	12639	0.954	ug/l	93
68) m/p-Xylenes	9.38	106	15001	2.949	ug/l	98
69) o-Xylene	9.78	106	10539	2.094	ug/l	98
80) 1,3,5-Trimethylbenzene	10.78	105	3505	0.303	ug/l	100
84) 1,2,4-Trimethylbenzene	11.15	105	8683	0.736	ug/l	98
86) p-Isopropyltoluene	11.48	119	59092	4.731	ug/l	95
95) Naphthalene	13.75	128	6688	1.195	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024551.D
Acq On : 14 Jun 2018 19:35
Operator : MD/SY
Sample : J3448-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
N-48969

Quant Time: Jun 15 05:00:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

