

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024562.D
 Acq On : 15 Jun 2018 00:36
 Operator : MD/SY
 Sample : J3452-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KY038PS027-180611

Quant Time: Jun 15 05:33:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	182987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308371	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	293775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	171424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	136424	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
35) Dibromofluoromethane	4.89	113	122201	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	7.57	98	457734	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	10.31	95	173129	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
Target Compounds						
16) Acetone	2.33	43	2601	1.620	ug/l	97
27) cis-1,2-Dichloroethene	4.24	96	110014	41.699	ug/l	85
29) Tetrahydrofuran	4.65	42	3369	2.523	ug/l #	88
44) Trichloroethene	6.19	130	18445	6.277	ug/l	94
64) Tetrachloroethene	8.23	164	27011	9.673	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024562.D
Acq On : 15 Jun 2018 00:36
Operator : MD/SY
Sample : J3452-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS027-180611

Quant Time: Jun 15 05:33:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

