

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024564.D
 Acq On : 15 Jun 2018 01:22
 Operator : MD/SY
 Sample : J3452-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KY038PS034-180611

Quant Time: Jun 15 05:36:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	187322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	307509	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	295639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	172867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	133769	43.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.48%	
35) Dibromofluoromethane	4.89	113	122901	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	7.57	98	458811	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	10.31	95	173605	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	15900	5.887	ug/l	Qvalue 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024564.D
Acq On : 15 Jun 2018 01:22
Operator : MD/SY
Sample : J3452-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS034-180611

Quant Time: Jun 15 05:36:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

