

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024567.D
 Acq On : 15 Jun 2018 02:31
 Operator : MD/SY
 Sample : J3478-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS050MW047-180610

Quant Time: Jun 15 05:42:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	171633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	284628	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	162928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	123909	44.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.44%	
35) Dibromofluoromethane	4.89	113	112658	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
50) Toluene-d8	7.57	98	425852	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	10.31	95	162793	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	15543	6.281	ug/l	80
44) Trichloroethene	6.19	130	1675	0.618	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024567.D
Acq On : 15 Jun 2018 02:31
Operator : MD/SY
Sample : J3478-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS050MW047-180610

Quant Time: Jun 15 05:42:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

