

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024571.D
 Acq On : 15 Jun 2018 04:04
 Operator : MD/SY
 Sample : J3478-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 KY105MW006-180611

Quant Time: Jun 15 05:51:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	183435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	302009	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	288635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	132370	44.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.40%	
35) Dibromofluoromethane	4.89	113	120620	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	7.57	98	442466	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	10.31	95	171465	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
Target Compounds						
4) Vinyl Chloride	1.41	62	4032	1.667	ug/l #	72
27) cis-1,2-Dichloroethene	4.23	96	1630	0.616	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024571.D
Acq On : 15 Jun 2018 04:04
Operator : MD/SY
Sample : J3478-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY105MW006-180611

Quant Time: Jun 15 05:51:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

