

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU061418\
 Data File : VU024577.D
 Acq On : 15 Jun 2018 06:22
 Operator : MD/SY
 Sample : J3478-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS050MW642-180612

Quant Time: Jun 15 08:35:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	169959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	278546	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	272190	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	162091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.32	65	119493	43.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.14%	
35) Dibromofluoromethane	4.89	113	111404	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	7.57	98	414053	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	10.31	95	159819	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	1736	0.708	ug/l	Qvalue 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU061418\
Data File : VU024577.D
Acq On : 15 Jun 2018 06:22
Operator : MD/SY
Sample : J3478-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS050MW642-180612

Quant Time: Jun 15 08:35:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

