

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061522\
 Data File : VU049180.D
 Acq On : 15 Jun 2022 19:07
 Operator : SY/MD
 Sample : N3285-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW4Q2

Quant Time: Jun 16 04:51:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:45:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.286	114	130493	5.000	ug/L	0.04
28) Chlorobenzene-d5	9.427	117	120841	5.000	ug/L	0.01
58) 1,4-Dichlorobenzene-d4	11.813	152	58771	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.611	65	42562	3.733	ug/L	0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%
7) Chloroethane-d5	1.932	69	32506	4.211	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.585	65	17363	3.786	ug/L	0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	4.726	46	142619	46.845	ug/L	0.09
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.700%
24) Chloroform-d	5.093	84	81099	4.533	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	5.745	65	46130	4.701	ug/L	0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	5.761	84	164866	4.529	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	6.723	67	57571	4.659	ug/L	0.03
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.916	98	145170	4.429	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.600%
43) trans-1,3-Dichloroprop...	8.195	79	22177	4.040	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	8.655	63	117251	52.953	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.900%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	47563	4.808	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	51999	4.611	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%
Target Compounds						Qvalue
14) Carbon disulfide	2.810	76	4256	0.176	ug/L	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061522\
Data File : VU049180.D
Acq On : 15 Jun 2022 19:07
Operator : SY/MD
Sample : N3285-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW4Q2

Quant Time: Jun 16 04:51:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:45:30 2022
Response via : Initial Calibration

