

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
 Data File : VU059280.D
 Acq On : 17 Jun 2024 15:12
 Operator : MD/SY
 Sample : P2895-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHB88

Quant Time: Jun 18 04:21:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 17 12:07:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.261	114	104258	5.000	ug/L	0.02
28) Chlorobenzene-d5	9.418	117	108128	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	48469	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	37563	5.181	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.600%
7) Chloroethane-d5	1.914	69	37231	5.337	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.800%
11) 1,1-Dichloroethene-d2	2.563	65	20212	6.720	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	134.400%#
20) 2-Butanone-d5	4.682	46	92055	61.645	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.280%
24) Chloroform-d	5.081	84	73751	4.956	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.718	65	38240	5.335	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.737	84	148373	4.986	ug/L	0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
36) 1,2-Dichloropropane-d6	6.698	67	47123	5.168	ug/L	0.02
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.400%
41) Toluene-d8	7.901	98	123752	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	8.184	79	13487	4.723	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.400%
46) 2-Hexanone-d5	8.640	63	65011	54.796	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.600%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	39978	5.306	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	45761	5.305	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.000%
Target Compounds						
					Qvalue	
13) Acetone	2.670	43	13130	13.554	ug/L	91
14) Carbon disulfide	2.782	76	50233	2.036	ug/L	99
16) Methylene chloride	3.042	84	1793	0.205	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
Data File : VU059280.D
Acq On : 17 Jun 2024 15:12
Operator : MD/SY
Sample : P2895-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHB88

Quant Time: Jun 18 04:21:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 17 12:07:46 2024
Response via : Initial Calibration

