




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
Data File : VU059290.D
Acq On : 17 Jun 2024 19:15
Operator : MD/SY
Sample : P2898-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD8

Manual IntegrationsAPPROVED

Quant Time: Jun 18 04:23:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 17 12:07:46 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 06/19/2024
Supervised By :Semsettin Yesilyurt 06/19/2024



(13) Acetone (T)

2.448min (-0.193) 168.17 ug/L

response 164360

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	2.48
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
Data File : VU059290.D
Acq On : 17 Jun 2024 19:15
Operator : MD/SY
Sample : P2898-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD8

Manual IntegrationsAPPROVED

Quant Time: Jun 18 04:23:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 17 12:07:46 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 06/19/2024
Supervised By :Semsettin Yesilyurt 06/19/2024



TIC: VU059290.D\data.ms

(13) Acetone (T)

2.621min (-0.019) 12.02 ug/L m

response 11750

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	34.67
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
 Data File : VU059290.D
 Acq On : 17 Jun 2024 19:15
 Operator : MD/SY
 Sample : P2898-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GD8

Manual IntegrationsAPPROVED

Quant Time: Jun 18 04:23:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 17 12:07:46 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 06/19/2024
 Supervised By :Semsettin Yesilyurt 06/19/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.245	114		105188	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117		108333	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152		48641	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.595	65		31294	4.279	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130		Recovery	=	85.600%	
7) Chloroethane-d5	1.907	69		32509	4.619	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130		Recovery	=	92.400%	
11) 1,1-Dichloroethene-d2	2.560	65		13286	4.378	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125		Recovery	=	87.600%	
20) 2-Butanone-d5	4.615	46		86166	57.191	ug/L	-0.01
Spiked Amount 50.000	Range 40	- 130		Recovery	=	114.380%	
24) Chloroform-d	5.058	84		69596	4.636	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	5.695	65		36423	5.036	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	100.800%	
32) Benzene-d6	5.721	84		134794	4.521	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125		Recovery	=	90.400%	
36) 1,2-Dichloropropane-d6	6.685	67		45242	4.952	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140		Recovery	=	99.000%	
41) Toluene-d8	7.894	98		109712	3.958	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130		Recovery	=	79.200%	
43) trans-1,3-Dichloroprop...	8.177	79		12868	4.498	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130		Recovery	=	90.000%	
46) 2-Hexanone-d5	8.627	63		61568	51.796	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130		Recovery	=	103.600%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84		37721	4.997	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120		Recovery	=	100.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152		43600	5.036	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120		Recovery	=	100.800%	
Target Compounds							
13) Acetone	2.621	43		11750m	12.023	ug/L	
16) Methylene chloride	3.039	84		3168	0.358	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061724\
Data File : VU059290.D
Acq On : 17 Jun 2024 19:15
Operator : MD/SY
Sample : P2898-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GD8

Manual IntegrationsAPPROVED

Quant Time: Jun 18 04:23:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 17 12:07:46 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 06/19/2024
Supervised By :Semsettin Yesilyurt 06/19/2024

