

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
 Data File : VU059318.D
 Acq On : 18 Jun 2024 17:58
 Operator : MD/SY
 Sample : P2856-19
 Misc : 4.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COSJ5

Quant Time: Jun 19 05:37:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 05:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	174460	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	175509	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	83057	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	37119	35.535	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.080%
7) Chloroethane-d5	1.869	69	19713	22.031	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	44.060%#
11) 1,1-Dichloroethene-d2	2.534	63	50277	24.385	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	48.760%#
21) 2-Butanone-d5	4.624	46	75601	76.786	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.790%
24) Chloroform-d	5.052	84	79475	38.219	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.440%
26) 1,2-Dichloroethane-d4	5.692	65	51733	40.590	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.180%
32) Benzene-d6	5.714	84	162221	40.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.540%
36) 1,2-Dichloropropane-d6	6.682	67	53564	38.565	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.140%
41) Toluene-d8	7.894	98	134704	39.392	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.780%#
43) trans-1,3-Dichloroprop...	8.177	79	20471	36.463	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.920%
47) 2-Hexanone-d5	8.637	63	46189	71.915	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.910%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	89561	38.101	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	76.200%
66) 1,2-Dichlorobenzene-d4	12.193	152	53184	42.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
Data File : VU059318.D
Acq On : 18 Jun 2024 17:58
Operator : MD/SY
Sample : P2856-19
Misc : 4.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SJ5

Quant Time: Jun 19 05:37:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 05:32:50 2024
Response via : Initial Calibration

