

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
 Data File : VU059332.D
 Acq On : 18 Jun 2024 23:25
 Operator : MD/SY
 Sample : P2873-03 200X
 Misc : 2.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0SG7

Quant Time: Jun 19 05:40:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 05:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	142821	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	144624	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	71199	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	37414	43.752	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	87.500%		
7) Chloroethane-d5	1.908	69	36896	50.369	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	100.740%		
11) 1,1-Dichloroethene-d2	2.560	63	49407	29.271	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	58.540%#		
21) 2-Butanone-d5	4.615	46	75015	93.069	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	93.070%		
24) Chloroform-d	5.055	84	81642	47.959	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.920%		
26) 1,2-Dichloroethane-d4	5.695	65	53103	50.895	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.800%		
32) Benzene-d6	5.721	84	158067	47.620	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.240%		
36) 1,2-Dichloropropane-d6	6.685	67	53300	46.571	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	93.140%		
41) Toluene-d8	7.894	98	126937	45.048	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.100%		
43) trans-1,3-Dichloroprop...	8.177	79	19535	42.226	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	84.460%		
47) 2-Hexanone-d5	8.628	63	41906	79.180	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.180%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	84376	43.561	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	87.120%		
66) 1,2-Dichlorobenzene-d4	12.190	152	53857	50.688	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.380%		
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	14925	10.930	ug/L	90
42) Toluene	7.968	91	22009	4.199	ug/L	99
46) Tetrachloroethene	8.547	164	168547	197.886	ug/L	91
52) Ethylbenzene	9.566	91	53317	9.866	ug/L	98
53) m,p-Xylene	9.689	106	23229	11.695	ug/L	96
54) o-Xylene	10.097	106	9377	4.871	ug/L	94
55) Styrene	10.116	104	6998	2.139	ug/L	93
62) 1,3,5-Trimethylbenzene	11.087	105	7267	2.002	ug/L	97
63) 1,2,4-Trimethylbenzene	11.467	105	20636	5.245	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
Data File : VU059332.D
Acq On : 18 Jun 2024 23:25
Operator : MD/SY
Sample : P2873-03 200X
Misc : 2.64g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SG7

Quant Time: Jun 19 05:40:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 05:32:50 2024
Response via : Initial Calibration

