

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
 Data File : VU059307.D
 Acq On : 18 Jun 2024 13:17
 Operator : MD/SY
 Sample : P2775-08ME
 Misc : 3.22g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4C86ME

Quant Time: Jun 19 04:40:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 03:23:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	163535	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	171287	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	73723	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	43021	43.937	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.880%
7) Chloroethane-d5	1.907	69	42208	50.322	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.640%
11) 1,1-Dichloroethene-d2	2.560	63	57113	29.551	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.100%#
21) 2-Butanone-d5	4.615	46	81515	88.323	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.320%
24) Chloroform-d	5.055	84	95492	48.990	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.980%
26) 1,2-Dichloroethane-d4	5.695	65	62696	52.478	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.960%
32) Benzene-d6	5.721	84	187939	47.806	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.620%
36) 1,2-Dichloropropane-d6	6.685	67	61895	45.662	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.320%
41) Toluene-d8	7.894	98	153375	45.958	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.920%
43) trans-1,3-Dichloroprop...	8.177	79	24540	44.788	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.580%
47) 2-Hexanone-d5	8.627	63	45013	71.811	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.810%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	98729	43.036	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.080%
66) 1,2-Dichlorobenzene-d4	12.190	152	58127	52.833	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
Data File : VU059307.D
Acq On : 18 Jun 2024 13:17
Operator : MD/SY
Sample : P2775-08ME
Misc : 3.22g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4C86ME

Quant Time: Jun 19 04:40:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 03:23:30 2024
Response via : Initial Calibration

