

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
 Data File : VU059309.D
 Acq On : 18 Jun 2024 14:04
 Operator : MD/SY
 Sample : P2775-16ME
 Misc : 3.11g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4CA1ME

Quant Time: Jun 19 04:40:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 03:23:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	157664	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	168145	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	73767	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	41981	44.471	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.940%
7) Chloroethane-d5	1.907	69	40935	50.622	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.240%
11) 1,1-Dichloroethene-d2	2.560	63	55780	29.936	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.880%#
21) 2-Butanone-d5	4.615	46	83283	93.599	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.600%
24) Chloroform-d	5.055	84	93472	49.739	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.480%
26) 1,2-Dichloroethane-d4	5.695	65	61896	53.738	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.480%
32) Benzene-d6	5.721	84	184884	47.907	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.820%
36) 1,2-Dichloropropane-d6	6.685	67	60863	45.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.480%
41) Toluene-d8	7.894	98	154124	47.045	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.080%
43) trans-1,3-Dichloroprop...	8.177	79	23752	44.160	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.320%
47) 2-Hexanone-d5	8.627	63	46209	75.097	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.100%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	97072	43.105	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	59120	53.704	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
Data File : VU059309.D
Acq On : 18 Jun 2024 14:04
Operator : MD/SY
Sample : P2775-16ME
Misc : 3.11g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4CA1ME

Quant Time: Jun 19 04:40:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 03:23:30 2024
Response via : Initial Calibration

