

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
 Data File : VU059316.D
 Acq On : 18 Jun 2024 17:12
 Operator : MD/SY
 Sample : P2856-21
 Misc : 2.71g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COSJ7

Quant Time: Jun 19 05:36:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 05:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	171128	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	174784	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	84438	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	35075	34.232	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.460%
7) Chloroethane-d5	1.869	69	19233	21.913	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	43.820%#
11) 1,1-Dichloroethene-d2	2.531	63	48514	23.988	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	47.980%#
21) 2-Butanone-d5	4.628	46	69831	72.306	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	72.310%
24) Chloroform-d	5.052	84	74917	36.729	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.460%
26) 1,2-Dichloroethane-d4	5.692	65	50630	40.498	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.000%
32) Benzene-d6	5.714	84	153982	38.385	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.760%
36) 1,2-Dichloropropane-d6	6.682	67	51100	36.944	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	73.880%
41) Toluene-d8	7.894	98	131518	38.620	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.240%#
43) trans-1,3-Dichloroprop...	8.174	79	21000	37.560	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.120%
47) 2-Hexanone-d5	8.644	63	43417	67.879	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	67.880%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	85366	36.467	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	72.940%
66) 1,2-Dichlorobenzene-d4	12.193	152	54068	42.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061824\
Data File : VU059316.D
Acq On : 18 Jun 2024 17:12
Operator : MD/SY
Sample : P2856-21
Misc : 2.71g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0SJ7

Quant Time: Jun 19 05:36:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 05:32:50 2024
Response via : Initial Calibration

