

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024762.D
 Acq On : 20 Jun 2018 17:07
 Operator : MD/SY
 Sample : J3577-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 540-IW-6

Quant Time: Jun 20 17:58:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305166	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	282496	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	156021	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	138190	42.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.84%	
35) Dibromofluoromethane	4.89	113	117547	46.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.82%	
50) Toluene-d8	7.57	98	443477	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
62) 4-Bromofluorobenzene	10.31	95	161637	44.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.14%	
Target Compounds						
16) Acetone	2.32	43	45850	26.50	ug/l	94
25) 2-Butanone	4.27	43	14013	6.03	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024762.D
 Acq On : 20 Jun 2018 17:07
 Operator : MD/SY
 Sample : J3577-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 540-IW-6

Quant Time: Jun 20 17:58:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

