

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024769.D
 Acq On : 20 Jun 2018 19:57
 Operator : MD/SY
 Sample : J3529-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TPTW-02

Quant Time: Jun 21 10:33:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	209982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	326324	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	306491	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	175317	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	145578	42.47	ug/l	0.00
Spiked Amount			Recovery	=	84.94%	
35) Dibromofluoromethane	4.89	113	127789	47.19	ug/l	0.00
Spiked Amount			Recovery	=	94.38%	
50) Toluene-d8	7.57	98	478750	48.46	ug/l	0.00
Spiked Amount			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	10.31	95	176205	44.93	ug/l	0.00
Spiked Amount			Recovery	=	89.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	2852	1.06	ug/l	100
11) Tert butyl alcohol	2.82	59	12615	15.13	ug/l	98
16) Acetone	2.32	43	110704	60.10	ug/l	95
19) Methyl tert-butyl Ether	3.00	73	33801	3.76	ug/l	98
25) 2-Butanone	4.27	43	27333	11.05	ug/l	98
40) Benzene	5.39	78	106586	9.75	ug/l	99
52) Toluene	7.64	92	7671	1.09	ug/l	98
67) Ethyl Benzene	9.25	91	125732	9.16	ug/l	97
68) m/p-Xylenes	9.38	106	25036	4.75	ug/l	98
69) o-Xylene	9.78	106	49380	9.47	ug/l	95
73) Isopropylbenzene	10.17	105	11722	0.87	ug/l	98
80) 1,3,5-Trimethylbenzene	10.78	105	18616	1.60	ug/l	100
84) 1,2,4-Trimethylbenzene	11.15	105	60317	5.07	ug/l	98
95) Naphthalene	13.75	128	3641431	279.32	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062018\
Data File : VU024769.D
Acq On : 20 Jun 2018 19:57
Operator : MD/SY
Sample : J3529-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TPTW-02

Quant Time: Jun 21 10:33:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

