

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024779.D
 Acq On : 21 Jun 2018 00:49
 Operator : MD/SY
 Sample : PB110408ZHE#20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB110408ZHE#20

Quant Time: Jun 21 08:40:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	195439	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	303868	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	282351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	152856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	137534	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	
35) Dibromofluoromethane	4.89	113	118458	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	7.57	98	438417	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
62) 4-Bromofluorobenzene	10.31	95	160041	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
Target Compounds						
16) Acetone	2.32	43	7267	4.24	ug/l	96
18) Methyl Acetate	2.62	43	16098	4.38	ug/l	97
20) Methylene Chloride	2.71	84	8795	3.27	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062018\
Data File : VU024779.D
Acq On : 21 Jun 2018 00:49
Operator : MD/SY
Sample : PB110408ZHE#20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB110408ZHE#20

Quant Time: Jun 21 08:40:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

