

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024781.D
 Acq On : 21 Jun 2018 01:37
 Operator : MD/SY
 Sample : J3245-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CL-02-061918-C

Quant Time: Jun 21 08:44:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	312130	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	289763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	160995	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	141989	43.63	ug/l	0.00
Spiked Amount			Recovery	=	87.26%	
35) Dibromofluoromethane	4.89	113	119652	46.19	ug/l	0.00
Spiked Amount			Recovery	=	92.38%	
50) Toluene-d8	7.57	98	448795	47.50	ug/l	0.00
Spiked Amount			Recovery	=	95.00%	
62) 4-Bromofluorobenzene	10.31	95	163795	43.66	ug/l	0.00
Spiked Amount			Recovery	=	87.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	9309	5.32	ug/l	97
18) Methyl Acetate	2.62	43	17275	4.61	ug/l	95
20) Methylene Chloride	2.71	84	275145	100.28	ug/l	97
68) m/p-Xylenes	9.38	106	10350	2.08	ug/l	99
84) 1,2,4-Trimethylbenzene	11.15	105	34324	3.14	ug/l	95
93) 1,2,4-Trichlorobenzene	13.51	180	5459	1.53	ug/l	91
95) Naphthalene	13.74	128	455072	38.60	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062018\
Data File : VU024781.D
Acq On : 21 Jun 2018 01:37
Operator : MD/SY
Sample : J3245-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CL-02-061918-C

Quant Time: Jun 21 08:44:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

