

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024792.D
 Acq On : 21 Jun 2018 06:05
 Operator : MD/SY
 Sample : J3585-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CKMW06

Quant Time: Jun 21 10:22:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	192358	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	298311	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	279296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	149110	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	137301	43.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.44%	
35) Dibromofluoromethane	4.89	113	114413	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	7.57	98	435332	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
62) 4-Bromofluorobenzene	10.31	95	158658	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	8.24	164	2477	0.93	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062018\
Data File : VU024792.D
Acq On : 21 Jun 2018 06:05
Operator : MD/SY
Sample : J3585-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CKMW06

Quant Time: Jun 21 10:22:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

