

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU062018\
 Data File : VU024794.D
 Acq On : 21 Jun 2018 06:53
 Operator : MD/SY
 Sample : J3585-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CKMW07

Quant Time: Jun 21 10:23:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200008	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305230	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	285105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	155317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	140060	42.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.80%	
35) Dibromofluoromethane	4.89	113	119888	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
50) Toluene-d8	7.57	98	446168	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	10.31	95	165220	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	4031	1.40	ug/l	83
44) Trichloroethene	6.19	130	2128	0.73	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU062018\
Data File : VU024794.D
Acq On : 21 Jun 2018 06:53
Operator : MD/SY
Sample : J3585-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CKMW07

Quant Time: Jun 21 10:23:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

